

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132200

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/28/2024	16:30
2	CAL2	1	Water	NA	NA	20.3	7.00	08/28/2024	16:31
3	CAL3	1	Water	NA	NA	20.3	10.02	08/28/2024	16:35
4	ICV	1	Water	NA	NA	20.4	7.01	08/28/2024	16:40
5	CCV1	1	Water	NA	NA	20.3	2.01	08/28/2024	16:45
6	P3772-01	1	Water	NA	NA	20.1	5.60	08/28/2024	16:50
7	P3772-01DUP	1	Water	NA	NA	20.2	5.61	08/28/2024	16:51
8	CCV2	1	Water	NA	NA	20.3	12.02	08/28/2024	17:00

WORKLIST(Hardcopy Internal Chain)

VB 132200

WorkList Name : ph3772 water

WorkList ID : 183043

Department : Wet-Chemistry

Date : 08-28-2024 16:10:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3772-01	-A BUR-1649	Water	pH	Cool 4 deg C	PSEG03	H51	08/28/2024	9040C

Date/Time 08/28/24 16:20

Raw Sample Received by: JO wsc

Raw Sample Relinquished by: JM (SM)

18:00

Date/Time 08/28/24

Raw Sample Received by: JM (SM)

Raw Sample Relinquished by: JO wsc