


LB132216
7470A INSTRUMENT ID: EV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.075	0.075	229	0.000	229	0				0.0
0.2	0.200	0.226	0.026	880	0.0 %	880					13
2.5	2.500	2.489	-0.011	10661	0.0 %	10661					0
5.0	5.000	4.882	-0.118	21001	0.0 %	21001					2
7.5	7.500	7.388	-0.112	31829	0.0 %	31829					-1
10.0	10.000	10.140	0.140	43719	0.0 %	43719					1

LB132216 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Aug 2024 11:41:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	229	-	0.0000	7470A	PPB	30 Aug 2024 11:55:23	S	Std
0.2 - 1	50.2	880	-	0.2000	7470A	PPB	30 Aug 2024 11:57:40	S	Std
2.5 - 1	52.5	10661	-	2.5000	7470A	PPB	30 Aug 2024 11:59:57	S	Std
5.0 - 1	55	21001	-	5.0000	7470A	PPB	30 Aug 2024 12:02:14	S	Std
7.5 - 1	59.5	31829	-	7.5000	7470A	PPB	30 Aug 2024 12:04:31	S	Std
10.0 - 1	5/8	43719	-	10.0000	7470A	PPB	30 Aug 2024 12:09:29	S	Std
ICV03 - 1	ICV03	17212	4.0053	-	7470A	PPB	30 Aug 2024 12:12:14	U	SMPL
ICB03 - 1	ICB03	-39	0.0130	-	7470A	PPB	30 Aug 2024 12:14:29	U	SMPL
CCV13 - 1	CCV13	21029	4.8887	-	7470A	PPB	30 Aug 2024 12:16:46	U	SMPL
CCB13 - 1	CCB13	-41	0.0125	-	7470A	PPB	30 Aug 2024 12:19:01	U	SMPL
CRA - 1	CRA	796	0.2062	-	7470A	PPB	30 Aug 2024 12:21:17	U	SMPL
HighStd - 1	HighStd	43282	10.0386	-	7470A	PPB	30 Aug 2024 12:23:32	U	SMPL
ChkStd - 1	ChkStd	29804	6.9195	-	7470A	PPB	30 Aug 2024 12:25:47	U	SMPL
PB163085BL - 1	PBW	-286	-0.0442	-	7470A	PPB	30 Aug 2024 12:28:05	U	SMPL
PB163085BS - 1	LCSW	16634	3.8716	-	7470A	PPB	30 Aug 2024 12:36:55	U	SMPL
P3726-02 - 1	MOO-SOIL-PILE	-178	-0.0192	-	7470A	PPB	30 Aug 2024 12:39:11	U	SMPL
P3726-02DUP - 1	MOO-SOIL-PILEDUP	-27	0.0158	-	7470A	PPB	30 Aug 2024 12:41:29	U	SMPL
P3726-02MS - 1	MOO-SOIL-PILEMS	16698	3.8864	-	7470A	PPB	30 Aug 2024 12:43:46	U	SMPL
P3726-02MSD - 1	MOO-SOIL-PILEMSD	16775	3.9042	-	7470A	PPB	30 Aug 2024 12:46:03	U	SMPL
P3727-04 - 1	MH-21	-186	-0.0210	-	7470A	PPB	30 Aug 2024 12:48:20	U	SMPL
CCV14 - 1	CCV14	20940	4.8681	-	7470A	PPB	30 Aug 2024 12:50:36	U	SMPL
CCB14 - 1	CCB14	79	0.0403	-	7470A	PPB	30 Aug 2024 12:52:52	U	SMPL
P3735-06 - 1	WC-11	82	0.0410	-	7470A	PPB	30 Aug 2024 12:55:09	U	SMPL
P3735-12 - 1	WC-13	216	0.0720	-	7470A	PPB	30 Aug 2024 12:57:24	U	SMPL
P3735-18 - 1	WC-16	111	0.0477	-	7470A	PPB	30 Aug 2024 12:59:40	U	SMPL
P3739-04 - 1	WC-4	617	0.1648	-	7470A	PPB	30 Aug 2024 13:01:55	U	SMPL
P3739-08 - 1	WC-1	-116	-0.0048	-	7470A	PPB	30 Aug 2024 13:04:11	U	SMPL
P3739-12 - 1	WC-2	-37	0.0134	-	7470A	PPB	30 Aug 2024 13:06:27	U	SMPL
P3739-16 - 1	WC-3	-3	0.0213	-	7470A	PPB	30 Aug 2024 13:08:43	U	SMPL
P3747-02 - 1	SLNP-DEMO	-31	0.0148	-	7470A	PPB	30 Aug 2024 13:11:00	U	SMPL
P3751-02 - 1	WC-CMC-ESLAG-1	-19	0.0176	-	7470A	PPB	30 Aug 2024 13:13:18	U	SMPL
P3758-02 - 1	SEW-13	178	0.0632	-	7470A	PPB	30 Aug 2024 13:15:33	U	SMPL
CCV15 - 1	CCV15	20292	4.7181	-	7470A	PPB	30 Aug 2024 13:17:49	U	SMPL
CCB15 - 1	CCB15	-82	0.0030	-	7470A	PPB	30 Aug 2024 13:20:04	U	SMPL
P3758-04 - 1	SEW-14	273	0.0852	-	7470A	PPB	30 Aug 2024 13:22:21	U	SMPL
P3760-01 - 1	SOIL-MOUND	5	0.0232	-	7470A	PPB	30 Aug 2024 13:24:37	U	SMPL
P3760-02 - 1	ASPHALT-MOUND	124	0.0507	-	7470A	PPB	30 Aug 2024 13:26:52	U	SMPL
PB162991TB - 1	PB162991TB	-23	0.0167	-	7470A	PPB	30 Aug 2024 13:29:07	U	SMPL
PB163048TB - 1	PB163048TB	-11	0.0195	-	7470A	PPB	30 Aug 2024 13:31:23	U	SMPL
P3726-02LX5 - 1		-18	0.0178	-	7470A	PPB	30 Aug 2024 13:33:40	U	SMPL
P3726-02A - 1		16654	3.8762	-	7470A	PPB	30 Aug 2024 13:35:56	U	SMPL
CCV16 - 1	CCV16	20541	4.7758	-	7470A	PPB	30 Aug 2024 13:38:13	U	SMPL
CCB16 - 1	CCB16	42	0.0317	-	7470A	PPB	30 Aug 2024 13:40:31	U	SMPL