

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132221

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/30/2024	09:10
2	CAL2	1	Water	NA	NA	20.2	7.01	08/30/2024	09:11
3	CAL3	1	Water	NA	NA	20.3	10.02	08/30/2024	09:12
4	ICV	1	Water	NA	NA	20.3	6.98	08/30/2024	09:15
5	CCV1	1	Water	NA	NA	20.3	2.01	08/30/2024	09:20
6	P3792-01	1	Water	NA	NA	20.2	4.16	08/30/2024	09:25
7	P3792-01DUP	1	Water	NA	NA	20.3	4.17	08/30/2024	09:26
8	P3792-02	1	Water	NA	NA	20.3	4.04	08/30/2024	09:30
9	P3792-03	1	Water	NA	NA	20.4	3.92	08/30/2024	09:35
10	P3792-04	1	Water	NA	NA	20.4	3.87	08/30/2024	09:37
11	P3795-01	1	Water	NA	NA	20.7	6.92	08/30/2024	09:44
12	P3799-07	1	Water	NA	NA	20.7	6.78	08/30/2024	09:47
13	P3799-08	1	Water	NA	NA	20.7	5.38	08/30/2024	09:50
14	P3799-09	1	Water	NA	NA	20.2	6.70	08/30/2024	10:00
15	CCV2	1	Water	NA	NA	20.3	12.02	08/30/2024	10:05