

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132242

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	09/03/2024	14:45
2	P3786-01		1	11	100.00	>212.0	>212.0	09/03/2024	16:10
3	P3786-01DUP		1	11	100.00	>212.0	>212.0	09/03/2024	16:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132242

WorkList Name : fp s-9-3

WorkList ID : 183119

Department : Wet-Chemistry

Date : 09-03-2024 08:56:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3786-01	0620	Solid	Flash Point	Cool 4 deg C	PSEG03	H53	08/29/2024	1010B

Date/Time 09/03/2024 15:10
 Raw Sample Received by: RM cwc
 Raw Sample Relinquished by: R (cwc)

Date/Time 09/03/2024 17:00
 Raw Sample Received by: R (cwc)
 Raw Sample Relinquished by: RM cwc