

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132243

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.4	09/04/2024	15:35
2	P3799-07		1	11	100.00	>212.0	>212.0	09/04/2024	16:00
3	P3799-07DUP		1	11	100.00	>212.0	>212.0	09/04/2024	16:30
4	P3799-08		1	13	100.00	>212.0	>212.0	09/04/2024	17:00
5	P3799-09		1	12	100.00	>212.0	>212.0	09/04/2024	17:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132243

WorkList Name : fp-9-3

WorkList ID : 183118

Department : Wet-Chemistry

Date : 09-03-2024 08:55:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3799-07	WELL-1	Water	Flash Point	Cool 4 deg C	PSEG03	H53	08/29/2024	1010B
P3799-08	WELL-2	Water	Flash Point	Cool 4 deg C	PSEG03	H53	08/29/2024	1010B
P3799-09	WELL-4	Water	Flash Point	Cool 4 deg C	PSEG03	H53	08/29/2024	1010B

Date/Time 02/04/2024 15:30
Raw Sample Received by: RM cws
Raw Sample Relinquished by: J. D. Miller

Date/Time 09/04/2024 17:40
Raw Sample Received by: MCO
Raw Sample Relinquished by: RM cws