


LB132249

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.0517e-004

C= -1.8725e-002

Rho= 0.9993272

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.051	0.051	338	0.000	338					-15
0.2	0.200	0.170	-0.030	919	0.0 %	919					-1
2.5	2.500	2.469	-0.031	12125	0.0 %	12125					-2
5.0	5.000	4.893	-0.107	23938	0.0 %	23938					-4
7.5	7.500	7.768	0.268	37950	0.0 %	37950					-2
10.0	10.000	9.850	-0.150	48102	0.0 %	48102					

LB132249 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Sep 2024 13:47:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	338	-	0.0000	7471B	PPB	03 Sep 2024 13:56:12	S	Std
0.2 - 1	50.2	919	-	0.2000	7471B	PPB	03 Sep 2024 13:58:28	S	Std
2.5 - 1	52.5	12125	-	2.5000	7471B	PPB	03 Sep 2024 14:00:45	S	Std
5.0 - 1	55	23938	-	5.0000	7471B	PPB	03 Sep 2024 14:03:02	S	Std
7.5 - 1	57.5	37950	-	7.5000	7471B	PPB	03 Sep 2024 14:05:21	S	Std
10.0 - 1	5N	48102	-	10.0000	7471B	PPB	03 Sep 2024 14:07:40	S	Std
ICV05 - 1	ICV05	19449	3.9716	-7471B	PPB	03 Sep 2024 14:10:28	U	SMPL	
ICB05 - 1	ICB05	100	0.0018	-7471B	PPB	03 Sep 2024 14:12:44	U	SMPL	
CCV21 - 1	CCV21	24198	4.9460	-7471B	PPB	03 Sep 2024 14:15:02	U	SMPL	
CCB21 - 1	CCB21	75	-0.0033	-7471B	PPB	03 Sep 2024 14:17:17	U	SMPL	
CRA - 1	CRA	924	0.1709	-7471B	PPB	03 Sep 2024 14:19:37	U	SMPL	
HighStd - 1	HighStd	48561	9.9446	-7471B	PPB	03 Sep 2024 14:21:52	U	SMPL	
ChkStd - 1	ChkStd	34395	7.0381	-7471B	PPB	03 Sep 2024 14:24:07	U	SMPL	
PB163113BL - 1	PBS	-292	-0.0786	-7471B	PPB	03 Sep 2024 14:26:30	U	SMPL	
PB163113BS - 1	LCSS	19515	3.9852	-7471B	PPB	03 Sep 2024 14:31:39	U	SMPL	
P3751-01 - 1	WC-CMC-ESLAG-1	25882	5.2915	-7471B	PPB	03 Sep 2024 14:33:55	U	SMPL	
P3751-01DUP - 1	WC-CMC-ESLAG-1DUP	21742	4.4421	-7471B	PPB	03 Sep 2024 14:36:14	U	SMPL	
P3751-01MS - 1	WC-CMC-ESLAG-1MS	42992	8.8020	-7471B	PPB	03 Sep 2024 14:38:35	U	SMPL	
P3751-01MSD - 1	WC-CMC-ESLAG-1MSD	43478	8.9017	-7471B	PPB	03 Sep 2024 14:40:55	U	SMPL	
P3771-01 - 1	DN-W-03	7650	1.5508	-7471B	PPB	03 Sep 2024 14:43:14	U	SMPL	
CCV22 - 1	CCV22	24047	4.9150	-7471B	PPB	03 Sep 2024 14:45:36	U	SMPL	
CCB22 - 1	CCB22	7	-0.0173	-7471B	PPB	03 Sep 2024 14:47:54	U	SMPL	
P3772-02 - 1	BUR-1294	39	-0.0107	-7471B	PPB	03 Sep 2024 14:50:13	U	SMPL	
P3773-02 - 1	AU-06-082824	1265	0.2408	-7471B	PPB	03 Sep 2024 14:52:28	U	SMPL	
P3787-01 - 1	SB-1	7809	1.5835	-7471B	PPB	03 Sep 2024 14:54:44	U	SMPL	
P3787-03 - 1	SB-2	8150	1.6534	-7471B	PPB	03 Sep 2024 14:57:00	U	SMPL	
P3787-05 - 1	SB-3	2754	0.5463	-7471B	PPB	03 Sep 2024 14:59:17	U	SMPL	
P3788-01 - 1	OR-02-082924	1789	0.3483	-7471B	PPB	03 Sep 2024 15:01:35	U	SMPL	
P3793-01 - 1	RT2864	1871	0.3652	-7471B	PPB	03 Sep 2024 15:03:53	U	SMPL	
P3794-01 - 1	CHRT-24376	17	-0.0152	-7471B	PPB	03 Sep 2024 15:06:10	U	SMPL	
P3794-03 - 1	RT-857	905	0.1670	-7471B	PPB	03 Sep 2024 15:08:26	U	SMPL	
P3795-03 - 1	239	664	0.1175	-7471B	PPB	03 Sep 2024 15:10:42	U	SMPL	
CCV23 - 1	CCV23	24895	5.0890	-7471B	PPB	03 Sep 2024 15:12:57	U	SMPL	
CCB23 - 1	CCB23	98	0.0014	-7471B	PPB	03 Sep 2024 15:15:12	U	SMPL	
P3797-01 - 1	207	3022	0.6013	-7471B	PPB	03 Sep 2024 15:17:32	U	SMPL	
P3799-01 - 1	TEST-PITS-1-2	994	0.1852	-7471B	PPB	03 Sep 2024 15:19:48	U	SMPL	
P3799-03 - 1	TEST-PITS-4	9509	1.9322	-7471B	PPB	03 Sep 2024 15:22:05	U	SMPL	
P3799-05 - 1	TEST-PITS-3	7913	1.6048	-7471B	PPB	03 Sep 2024 15:24:27	U	SMPL	
P3811-02 - 1	DRUM-2	53	-0.0079	-7471B	PPB	03 Sep 2024 15:26:48	U	SMPL	
P3812-01 - 1	EO-02-083024	2119	0.4160	-7471B	PPB	03 Sep 2024 15:29:06	U	SMPL	
P3813-01 - 1	TR-04-083024	2719	0.5391	-7471B	PPB	03 Sep 2024 15:31:22	U	SMPL	
P3813-03 - 1	TR-06-083024	2549	0.5043	-7471B	PPB	03 Sep 2024 15:33:39	U	SMPL	
P3751-01LX5 - 1		4892	0.9850	-7471B	PPB	03 Sep 2024 15:35:57	U	SMPL	
P3751-01A - 1		43110	8.8262	-7471B	PPB	03 Sep 2024 15:38:13	U	SMPL	
CCV24 - 1	CCV24	24544	5.0170	-7471B	PPB	03 Sep 2024 15:40:30	U	SMPL	
CCB24 - 1	CCB24	68	-0.0048	-7471B	PPB	03 Sep 2024 15:42:51	U	SMPL	