

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132253

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/03/2024	15:20
2	CAL2	1	Water	NA	NA	20.3	7.00	09/03/2024	15:21
3	CAL3	1	Water	NA	NA	20.3	10.02	09/03/2024	15:25
4	ICV	1	Water	NA	NA	20.3	7.00	09/03/2024	15:30
5	CCV1	1	Water	NA	NA	20.3	2.01	09/03/2024	15:33
6	P3820-03	1	Water	NA	NA	20.7	6.91	09/03/2024	15:40
7	P3820-03DUP	1	Water	NA	NA	20.8	6.92	09/03/2024	15:41
8	P3821-01	1	Water	NA	NA	24.6	7.37	09/03/2024	15:47
9	CCV2	1	Water	NA	NA	20.3	12.02	09/03/2024	15:50

WORKLIST(Hardcopy Internal Chain)

VB 132253

WorkList Name : PH P3821 WorkList ID : 183139 Department : Wet-Chemistry Date : 09-03-2024 15:09:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3820-03	WATER-1	Water	pH	Cool 4 deg C	PSEG03	H51	09/03/2024	9040C
P3821-01	MANHOLE-5A-WATER	Water	pH	Cool 4 deg C	PSEG03	H51	09/03/2024	9040C

Date/Time 09/03/24 15:15
Raw Sample Received by: JG awc
Raw Sample Relinquished by: JT (SM)

Date/Time 09/03/24 17:00
Raw Sample Received by: JT (SM)
Raw Sample Relinquished by: JG awc