

TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 09/03/2024 15:35 **TEMP1 OUT:** 103 °C 09/04/2024 07:30
TEMP2 IN: 104 °C 09/04/2024 08:00 **TEMP2 OUT:** 104 °C 09/04/2024 09:30
TEMP3 IN: 540 °C 09/04/2024 10:00 **TEMP3 OUT:** 540 °C 09/04/2024 11:30
TEMP4 IN: 560 °C 09/04/2024 12:00 **TEMP4 OUT:** 560 °C 09/04/2024 13:30

Run Number: LB132255

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB132255BL	65.0893	65.0893	100	65.0893	65.0893	65.0893	65.0893	0.0000	0
2	P3821-01	88.9530	88.9530	100	89.3024	89.3024	88.9620	88.9620	0.3404	3404
3	P3821-01DUP	91.6429	91.6429	100	91.9925	91.9925	91.6609	91.6609	0.3316	3316

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

LB 132255

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TVS P3821 WATER

WorkList ID : 183141

Department : Wet-Chemistry

Date : 09-03-2024 15:11:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3821-01	MANHOLE-5A-WATER	Water	TVS	Cool 4 deg C	PSEG03	H51	09/03/2024	SM2540 G

Date/Time09/03/2415:15

Raw Sample Received by: JG WJC

Raw Sample Relinquished by: JT (in)

Date/Time09/03/2417:00

Raw Sample Received by: JT (in)

Raw Sample Relinquished by: JG WJC