

LB132263


SFAM01.1

INSTRUMENT 11: CV1

Linear

 μ Abs.:

40954



A= 0.0000e+000

B= 2.4574e-004 slope

C= -1.5746e-001 $\rightarrow$ intercept

Rho= 0.9998313

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.000	-0.000	639	0.000	639					1
0.05	0.050					442	0				1
0.20	0.200	0.198	-0.002	1447	0.0 %	1447					1
2.50	2.500	2.460	-0.040	10652	0.0 %	10652					2
5.00	5.000	5.004	0.004	21004	0.0 %	21004					0
7.50	7.500	7.631	0.131	31695	0.0 %	31695					2
10.0	10.000	9.907	-0.093	40954	0.0 %	40954					1

LB132263

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	639	-		0 SFAM01.1	PPB	9/4/2024 8:01	Std
	0.2 S01	1447	-		0.2 SFAM01.1	PPB	9/4/2024 8:03	Std
	2.5 S02	10652	-		2.5 SFAM01.1	PPB	9/4/2024 8:05	Std
	5 S03	21004	-		5 SFAM01.1	PPB	9/4/2024 8:07	Std
	7.5 S04	31695	-		7.5 SFAM01.1	PPB	9/4/2024 8:10	Std
	10 S05	40954	-		10 SFAM01.1	PPB	9/4/2024 8:15	Std
ICV006	ICV006	17034	4.0285 -		SFAM01.1	PPB	9/4/2024 8:18	SMPL
ICB006	ICB006	489	-0.0373 -		SFAM01.1	PPB	9/4/2024 8:20	SMPL
CCV025	CCV025	21363	5.0923 -		SFAM01.1	PPB	9/4/2024 8:25	SMPL
CCB025	CCB025	420	-0.0543 -		SFAM01.1	PPB	9/4/2024 8:27	SMPL
PB163135BL	PBW135	495	-0.0358 -		SFAM01.1	PPB	9/4/2024 8:29	SMPL
P3584-01	MDCUY6	566	-0.0184 -		SFAM01.1	PPB	9/4/2024 8:32	SMPL
P3584-02	MDCUY9	685	0.0109 -		SFAM01.1	PPB	9/4/2024 8:34	SMPL
P3584-03	MDCUY0	665	0.006 -		SFAM01.1	PPB	9/4/2024 8:36	SMPL
P3584-04	MDCUY0D	635	-0.0014 -		SFAM01.1	PPB	9/4/2024 8:38	SMPL
P3584-05	MDCUY0S	4894	1.0452 -		SFAM01.1	PPB	9/4/2024 8:41	SMPL
P3584-06	MDCUZ2	570	-0.0174 -		SFAM01.1	PPB	9/4/2024 8:43	SMPL
P3584-07	MDCUZ5	616	-0.0061 -		SFAM01.1	PPB	9/4/2024 8:45	SMPL
P3584-08	MDCVH8	582	-0.0144 -		SFAM01.1	PPB	9/4/2024 8:47	SMPL
P3584-09	MDCVN6	526	-0.0282 -		SFAM01.1	PPB	9/4/2024 8:50	SMPL
P3584-10	MDCVN8	542	-0.0243 -		SFAM01.1	PPB	9/4/2024 8:52	SMPL
P3585-01	MDCUY8	583	-0.0142 -		SFAM01.1	PPB	9/4/2024 8:54	SMPL
P3585-02	MDCUZ1	610	-0.0076 -		SFAM01.1	PPB	9/4/2024 8:56	SMPL
P3585-03	MDCUY2	729	0.0217 -		SFAM01.1	PPB	9/4/2024 8:59	SMPL
P3585-04	MDCUY2D	666	0.0062 -		SFAM01.1	PPB	9/4/2024 9:01	SMPL
P3585-05	MDCUY2S	4900	1.0467 -		SFAM01.1	PPB	9/4/2024 9:03	SMPL
P3585-06	MDCUZ4	624	-0.0041 -		SFAM01.1	PPB	9/4/2024 9:05	SMPL
P3585-07	MDCUZ7	605	-0.0088 -		SFAM01.1	PPB	9/4/2024 9:08	SMPL
P3585-08	MDCVH9	592	-0.012 -		SFAM01.1	PPB	9/4/2024 9:10	SMPL
P3585-09	MDCVN7	506	-0.0331 -		SFAM01.1	PPB	9/4/2024 9:12	SMPL
CCV026	CCV026	20767	4.9459 -		SFAM01.1	PPB	9/4/2024 9:15	SMPL
CCB026	CCB026	455	-0.0457 -		SFAM01.1	PPB	9/4/2024 9:17	SMPL
P3585-10	MDCVN9	598	-0.0105 -		SFAM01.1	PPB	9/4/2024 9:19	SMPL
CCV027	CCV027	21104	5.0287 -		SFAM01.1	PPB	9/4/2024 9:21	SMPL
CCB027	CCB027	557	-0.0206 -		SFAM01.1	PPB	9/4/2024 9:24	SMPL