

LB132269


7470A

INSTRUMENT ID: CY1

Linear

 μ Abs.:

33245



A= 0.0000e+000

B= 3.0616e-004

C= -1.6713e-001

Rho= 0.9990670

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/P D
0.0	0.000	0.000	0.000	547	0.000	547	0				0.00
0.2	0.200	0.231	0.031	1300	0.0 %	1300					16
2.5	2.500	2.657	0.157	9223	0.0 %	9223					6
5.0	5.000	4.669	-0.331	15796	0.0 %	15796					-7
7.5	7.500	7.631	0.131	25471	0.0 %	25471					2
10.0	10.000	10.012	0.012	33249	0.0 %	33249					0

LB132269 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Sep 2024 12:26:15

Sample ID	Extended ID	Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	547	-	0.0000	7470A	PPB	04 Sep 2024 12:26:25	S	Std
0.2 - 1	51.2	1300	-	0.2000	7470A	PPB	04 Sep 2024 12:28:42	S	Std
2.5 - 1	52.5	9223	-	2.5000	7470A	PPB	04 Sep 2024 12:30:59	S	Std
5.0 - 1	55	15796	-	5.0000	7470A	PPB	04 Sep 2024 12:35:51	S	Std
7.5 - 1	57.5	25471	-	7.5000	7470A	PPB	04 Sep 2024 12:41:23	S	Std
10.0 - 1	510	33249	-	10.0000	7470A	PPB	04 Sep 2024 12:46:20	S	Std
ICV08 - 1	ICV08	12792	3.7492	-	7470A	PPB	04 Sep 2024 12:54:13	U	SMPL
ICB08 - 1	ICB08	275	-0.0829	-	7470A	PPB	04 Sep 2024 12:56:29	U	SMPL
CCV31 - 1	CCV31	15949	4.7158	-	7470A	PPB	04 Sep 2024 13:01:31	U	SMPL
CCB31 - 1	CCB31	253	-0.0897	-	7470A	PPB	04 Sep 2024 13:03:47	U	SMPL
CRA - 1	CRA	1220	0.2064	-	7470A	PPB	04 Sep 2024 13:06:04	U	SMPL
HighStd - 1	HighStd	32619	9.8194	-	7470A	PPB	04 Sep 2024 13:08:18	U	SMPL
ChkStd - 1	ChkStd	22413	6.6948	-	7470A	PPB	04 Sep 2024 13:10:33	U	SMPL
PB163151BL - 1	PBW	720	0.0533	-	7470A	PPB	04 Sep 2024 13:15:28	U	SMPL
PB163151BS - 1	LCSW	12579	3.6840	-	7470A	PPB	04 Sep 2024 13:17:44	U	SMPL
P3609-01 - 1	915-J-WS-081424	222	-0.0992	-	7470A	PPB	04 Sep 2024 13:19:59	U	SMPL
P3609-01DUP - 1	915-J-WS-081424DUP	275	-0.0829	-	7470A	PPB	04 Sep 2024 13:24:54	U	SMPL
P3609-01MS - 1	915-J-WS-081424MS	16822	4.9830	-	7470A	PPB	04 Sep 2024 13:27:10	U	SMPL
P3609-01MSD - 1	915-J-WS-081424MSD	16335	4.8339	-	7470A	PPB	04 Sep 2024 13:29:27	U	SMPL
P3609-02 - 1	920-J-WS-081424	-90	-0.1947	-	7470A	PPB	04 Sep 2024 13:31:45	U	SMPL
CCV32 - 1	CCV32	16021	4.7378	-	7470A	PPB	04 Sep 2024 13:34:03	U	SMPL
CCB32 - 1	CCB32	471	-0.0229	-	7470A	PPB	04 Sep 2024 13:36:18	U	SMPL
P3609-01LX5 - 1		8	-0.1647	-	7470A	PPB	04 Sep 2024 13:38:35	U	SMPL
P3609-01A - 1		11967	3.4967	-	7470A	PPB	04 Sep 2024 13:40:50	U	SMPL
CCV33 - 1	CCV33	16809	4.9791	-	7470A	PPB	04 Sep 2024 13:43:06	U	SMPL
CCB33 - 1	CCB33	306	-0.0734	-	7470A	PPB	04 Sep 2024 13:45:24	U	SMPL