

Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB132272
Analysis Date: 09/04/2024
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 09/04/2024
Extraction IN Time: 13:50
Extraction OUT Time: 14:25
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB132272BL	LB132272BL	WATER	1.3	1000	100	3.0526	3.0526	3.01	3.0527	3.0527	0.0001	0.1
2	LB132272BS	LB132272BS	WATER	1.3	1000	100	3.1974	3.1974	3.02	3.2142	3.2142	0.0168	16.8
3	LB132272BSD	LB132272BSD	WATER	1.3	1000	100	3.0877	3.0877	3.03	3.1048	3.1048	0.0171	17.1
4	P3821-01	MANHOLE-5A-WATER	WATER	1.3	1000	100	3.0444	3.0444	3.02	3.0447	3.0447	0.0003	0.3
5	P3823-01	402	WATER	1.3	1000	100	3.1184	3.1184	3.01	3.1187	3.1187	0.0003	0.3
6	P3833-01	GRAB	WATER	1.6	1000	100	3.0460	3.0460	3.04	3.0589	3.0589	0.0129	12.9



QC Batch# LB132272 Test: TPH Analysis Date: 09/04/2024

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3110
pH Paper 0-14	M6069
Sodium Sulfate	EP2532
1:1 HCL	WP108566
Silica Gel	W3079
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP108567
LCSWD	5.00 ML	WP108568
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance:	<u>0.0018</u> (0.0018-0.0022)	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1	<u>15:41</u>
1.0000 gram Balance:	<u>1.0005</u> (0.9950-1.0050)	In Time1:	<u>15:00</u>		
Bal Check Time:	<u>14:15</u>	Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1	<u>16:15</u>
		Out Time1	<u>15:40</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u> (0.0018-0.0022)	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2	<u>17:31</u>
1.0000 gram Balance:	<u>1.0004</u> (0.9950-1.0050)	In Time2:	<u>16:50</u>		
Bal Check Time:	<u>18:11</u>	Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2	<u>18:10</u>
		Out Time2:	<u>17:30</u>		

WB132272

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tph p3833 WorkList ID : 183167 Department : Wet-Chemistry Date : 09-04-2024 13:36:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3821-01	MANHOLE-5A-WATER	Water	TPH	Conc H2SO4 to pH < 2	PSEG03	H51	09/03/2024	1664A
P3823-01	402	Water	TPH	Conc H2SO4 to pH < 2	PSEG04	H52	09/03/2024	1664A
P3833-01	GRAB	Water	TPH	Conc H2SO4 to pH < 2	ARAM01	H53	09/04/2024	1664A

Date/Time 09/04/24 13:45
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 09/04/24 17:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]