

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132276

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/04/2024	16:25
2	CAL2	1	Water	NA	NA	20.3	7.01	09/04/2024	16:26
3	CAL3	1	Water	NA	NA	20.3	10.02	09/04/2024	16:29
4	ICV	1	Water	NA	NA	20.3	7.00	09/04/2024	16:33
5	CCV1	1	Water	NA	NA	20.3	2.01	09/04/2024	16:35
6	P3837-02	1	Water	NA	NA	23.6	7.01	09/04/2024	16:40
7	P3837-02DUP	1	Water	NA	NA	23.7	7.02	09/04/2024	16:41
8	CCV2	1	Water	NA	NA	20.3	12.02	09/04/2024	16:45

WORKLIST(Hardcopy Internal Chain)

VB132276

WorkList Name : ph p3837

WorkList ID : 183170

Department : Wet-Chemistry

Date : 09-04-2024 16:15:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3837-02	C	RAIN-TANK	Water	pH	Cool 4 deg C	PSEG03	H53	09/04/2024 9040C

Date/Time 09/04/24 16:20
Raw Sample Received by: Jd wee
Raw Sample Relinquished by: Jc (su)

Date/Time 09/04/24 17:30
Raw Sample Received by: Jc (su)
Raw Sample Relinquished by: Jd wee