


LB132284

SFAM01.1

INSTRUMENT ID: CY1



Linear

A= 0.0000e+000
 B= 2.6017e-004 slope
 C= -1.9834e-003 -121mV/deg
 Rho= 0.9999044
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/h
0.00	0.000	0.027	0.027	110	0.000	110					
0.05	0.050					442	0				
0.20	0.200	0.209	0.009	811	0.0 %	811					
2.50	2.500	2.447	-0.053	9413	0.0 %	9413					
5.00	5.000	4.964	-0.036	19089	0.0 %	19089					
7.50	7.500	7.595	0.095	29200	0.0 %	29200					
10.0	10.000	9.958	-0.042	38284	0.0 %	38284					

o/h
 11
 12
 11
 11
 0

LB132284

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	110	-		0 SFAM01.1	PPB	9/5/2024 10:06	Std
	0.2 S01	811	-		0.2 SFAM01.1	PPB	9/5/2024 10:09	Std
	2.5 S02	9413	-		2.5 SFAM01.1	PPB	9/5/2024 10:13	Std
	5 S03	19089	-		5 SFAM01.1	PPB	9/5/2024 10:15	Std
	7.5 S04	29200	-		7.5 SFAM01.1	PPB	9/5/2024 10:17	Std
	10 S05	38284	-		10 SFAM01.1	PPB	9/5/2024 10:20	Std
ICV009	ICV009	15433	4.0132 -		SFAM01.1	PPB	9/5/2024 10:23	SMPL
ICB009	ICB009	-219	-0.059 -		SFAM01.1	PPB	9/5/2024 10:25	SMPL
CCV034	CCV034	19395	5.0439 -		SFAM01.1	PPB	9/5/2024 10:27	SMPL
CCB034	CCB034	-65	-0.0189 -		SFAM01.1	PPB	9/5/2024 10:29	SMPL
PB163163BL	PBS163	50	0.011 -		SFAM01.1	PPB	9/5/2024 10:32	SMPL
P3649-01	MDCUZ8	2725	0.707 -		SFAM01.1	PPB	9/5/2024 10:39	SMPL
P3649-02	MDCUZ8D	2278	0.5907 -		SFAM01.1	PPB	9/5/2024 10:41	SMPL
P3649-03	MDCUZ8S	13117	3.4106 -		SFAM01.1	PPB	9/5/2024 10:43	SMPL
P3649-04	MDCVA0	6878	1.7874 -		SFAM01.1	PPB	9/5/2024 10:46	SMPL
P3649-05	MDCVA1	1059	0.2735 -		SFAM01.1	PPB	9/5/2024 10:48	SMPL
P3649-06	MDCVA2	1362	0.3524 -		SFAM01.1	PPB	9/5/2024 10:50	SMPL
P3649-07	MDCVA3	564	0.1448 -		SFAM01.1	PPB	9/5/2024 10:53	SMPL
P3649-08	MDCVA4	742	0.1911 -		SFAM01.1	PPB	9/5/2024 10:55	SMPL
P3649-09	MDCVO4	566	0.1453 -		SFAM01.1	PPB	9/5/2024 10:57	SMPL
CCV035	CCV035	18950	4.9282 -		SFAM01.1	PPB	9/5/2024 10:59	SMPL
CCB035	CCB035	-205	-0.0553 -		SFAM01.1	PPB	9/5/2024 11:02	SMPL