

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132291

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/06/2024	08:05
2	CAL2	1	Water	NA	NA	20.2	7.00	09/06/2024	08:06
3	CAL3	1	Water	NA	NA	20.2	10.02	09/06/2024	08:08
4	ICV	1	Water	NA	NA	20.3	7.01	09/06/2024	08:10
5	CCV1	1	Water	NA	NA	20.3	2.01	09/06/2024	08:11
6	P3850-01	1	Water	NA	NA	20.3	5.08	09/06/2024	08:15
7	P3852-06	1	Water	NA	NA	22.3	9.03	09/06/2024	08:20
8	P3853-01	1	Water	NA	NA	21.5	7.45	09/06/2024	08:27
9	P3853-01DUP	1	Water	NA	NA	21.6	7.46	09/06/2024	08:28
10	CCV2	1	Water	NA	NA	20.3	12.02	09/06/2024	08:30

WORKLIST(Hardcopy Internal Chain)

132291

WorkList Name : ph p3852 water WorkList ID : 183200 Department : Wet-Chemistry Date : 09-06-2024 07:47:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3850-01	50428	Water	pH	Cool 4 deg C	PSEG03	H53	09/05/2024	9040C
P3852-06	D3659	Water	pH	Cool 4 deg C	PSEG03	H53	09/05/2024	9040C
P3853-01	252805	Water	pH	Cool 4 deg C	PSEG03	H53	09/05/2024	9040C

Date/Time 09/06/24 08:00
Raw Sample Received by: Jm wee!
Raw Sample Relinquished by: Jm (SM)

Date/Time 09/06/24 11:30
Raw Sample Received by: Jm (SM)
Raw Sample Relinquished by: Jm wee!