

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB132292

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/06/2024	08:05
2	CAL2	1	Water	NA	NA	20.2	7.00	09/06/2024	08:06
3	CAL3	1	Water	NA	NA	20.2	10.02	09/06/2024	08:08
4	ICV	1	Water	NA	NA	20.3	7.01	09/06/2024	08:10
5	CCV1	1	Water	NA	NA	20.3	2.02	09/06/2024	08:35
6	P3852-01	1	Solid	20.02	20	24.6	7.92	09/06/2024	08:37
7	P3857-02	1	Solid	20.03	20	23.8	6.94	09/06/2024	08:40
8	P3857-02DUP	1	Solid	20.04	20	23.9	6.95	09/06/2024	08:41
9	P3857-05	1	Solid	20.01	20	24.1	7.02	09/06/2024	08:45
10	P3857-08	1	Solid	20.03	20	23.9	8.07	09/06/2024	08:49
11	P3857-11	1	Solid	20.02	20	23.6	8.00	09/06/2024	08:50
12	CCV2	1	Water	NA	NA	20.8	12.02	09/06/2024	09:00

WORKLIST(Hardcopy Internal Chain)

W132292

WorkList Name : corrosivity p3857

WorkList ID : 183201

Department : Wet-Chemistry

Date : 09-06-2024 07:49:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3852-01	D3656	Solid	Corrosivity	Cool 4 deg C	PSEG03	H53	09/05/2024	9045D
P3857-02	WC-C1-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	H53	09/05/2024	9045D
P3857-05	WC-C2-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	H53	09/05/2024	9045D
P3857-08	WC-D1-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	H53	09/05/2024	9045D
P3857-11	WC-D2-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	H53	09/05/2024	9045D

Date/Time 09/06/24 08:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time

09/06/24

Raw Sample Received by:

Raw Sample Relinquished by:

11:00

[Signature]

[Signature]