

LB132294


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

3851E



A= 0.0000e+000

B= 2.5992e-004

C= 3.0105e-002

Rho= 0.9993728

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.0	0.000	0.044	0.044	52	0.000	52	0				13
0.2	0.200	0.234	0.034	785	0.0 %	785					13
2.5	2.500	2.322	-0.178	8817	0.0 %	8817					13
5.0	5.000	5.208	0.208	19923	0.0 %	19923					4
7.5	7.500	7.350	-0.150	28163	0.0 %	28163					2
10.0	10.000	10.042	0.042	38518	0.0 %	38518					0

LB132294 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Sep 2024 08:35:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	52	-	0.0000	7470A	PPB	06 Sep 2024 08:49:59	S	Std
0.2 - 1	50.2	785	-	0.2000	7470A	PPB	06 Sep 2024 08:52:15	S	Std
2.5 - 1	52.5	8817	-	2.5000	7470A	PPB	06 Sep 2024 08:54:32	S	Std
5.0 - 1	55	19923	-	5.0000	7470A	PPB	06 Sep 2024 08:56:49	S	Std
7.5 - 1	57.5	28163	-	7.5000	7470A	PPB	06 Sep 2024 08:59:05	S	Std
10.0 - 1	510	38518	-	10.0000	7470A	PPB	06 Sep 2024 09:04:04	S	Std
ICV10 - 1	ICV10	15201	3.9811	-	7470A	PPB	06 Sep 2024 09:08:17	U	SMPL
ICB10 - 1	ICB10	-246	-0.0338	-	7470A	PPB	06 Sep 2024 09:10:32	U	SMPL
CCV36 - 1	CCV36	20550	5.3715	-	7470A	PPB	06 Sep 2024 09:12:48	U	SMPL
CCB36 - 1	CCB36	-452	-0.0874	-	7470A	PPB	06 Sep 2024 09:15:03	U	SMPL
CRA - 1	CRA	586	0.1824	-	7470A	PPB	06 Sep 2024 09:17:20	U	SMPL
HighStd - 1	HighStd	38392	10.0090	-	7470A	PPB	06 Sep 2024 09:19:34	U	SMPL
ChkStd - 1	ChkStd	26443	6.9032	-	7470A	PPB	06 Sep 2024 09:21:50	U	SMPL
PB163172BL - 1	PBW	-442	-0.0848	-	7470A	PPB	06 Sep 2024 09:24:07	U	SMPL
PB163172BS - 1	LCSW	16777	4.3908	-	7470A	PPB	06 Sep 2024 09:26:25	U	SMPL
P3820-03 - 1	WATER-1	-231	-0.0299	-	7470A	PPB	06 Sep 2024 09:28:41	U	SMPL
P3821-01 - 1	MANHOLE-5A-WATER	-14	0.0265	-	7470A	PPB	06 Sep 2024 09:30:59	U	SMPL
P3833-02 - 1	COMP	283	0.1037	-	7470A	PPB	06 Sep 2024 09:33:16	U	SMPL
P3837-01 - 1	902	865	0.2549	-	7470A	PPB	06 Sep 2024 09:35:33	U	SMPL
P3837-02 - 1	RAIN-TANK	634	0.1949	-	7470A	PPB	06 Sep 2024 09:37:48	U	SMPL
CCV37 - 1	CCV37	19277	5.0406	-	7470A	PPB	06 Sep 2024 09:40:06	U	SMPL
CCB37 - 1	CCB37	-399	-0.0736	-	7470A	PPB	06 Sep 2024 09:42:21	U	SMPL
P3839-03 - 1	829	174	0.0753	-	7470A	PPB	06 Sep 2024 09:58:55	U	SMPL
P3839-03DUP - 1	829DUP	91	0.0538	-	7470A	PPB	06 Sep 2024 10:01:09	U	SMPL
P3839-03MS - 1	829MS	16454	4.3068	-	7470A	PPB	06 Sep 2024 10:03:25	U	SMPL
P3839-03MSD - 1	829MSD	15514	4.0625	-	7470A	PPB	06 Sep 2024 10:05:40	U	SMPL
P3839-04 - 1	831	1353	0.3818	-	7470A	PPB	06 Sep 2024 10:08:02	U	SMPL
P3853-01 - 1	252805	1687	0.4686	-	7470A	PPB	06 Sep 2024 10:10:22	U	SMPL
P3839-03LX5 - 1		-85	0.0080	-	7470A	PPB	06 Sep 2024 10:12:38	U	SMPL
P3839-03A - 1		15668	4.1025	-	7470A	PPB	06 Sep 2024 10:14:54	U	SMPL
CCV38 - 1	CCV38	21030	5.4962	-	7470A	PPB	06 Sep 2024 10:17:11	U	SMPL
CCB38 - 1	CCB38	-193	-0.0201	-	7470A	PPB	06 Sep 2024 10:19:29	U	SMPL