

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Niha

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB132300

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP109735
Chlorine Calibration std, 0.1ppm	WP109730
Chlorine Calibration std, 0.2ppm	WP109731
Chlorine Calibration std, 0.8ppm	WP109732
Chlorine Calibration std, 0.0ppm	WP109729
Chlorine Calibration std, 1.6ppm	WP109733
Residual Chlorine Calibration and CCV std, 0	WP109734
Total Chlorine Powder Pillows	W2815

Intercept: -0.0057

Slope: 0.9982

Regression: 0.999854

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	0.01		09/16/2024	16:00
2	CAL2	0.1	1	0.000	0.100	0.100	0.11	6	09/16/2024	16:03
3	CAL3	0.2	1	0.000	0.200	0.200	0.21	3	09/16/2024	16:06
4	CAL4	0.4	1	0.000	0.380	0.380	0.39	-3.5	09/16/2024	16:09
5	CAL5	0.8	1	0.000	0.780	0.780	0.79	-1.6	09/16/2024	16:12
6	CAL6	1.6	1	0.000	1.600	1.600	1.61	0.6	09/16/2024	16:15

Analysis Method: SM4500 Cl GANALYST: NihaParameter: Residual ChlorineSUPERVISOR REVIEW BY: IwonaRun Number: LB132300

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4000	0.4000	0.4060	09/16/2024	16:18
2	ICB				1	0.0000	0.0000	0.0000	0.0060	09/16/2024	16:21
3	CCV1			0.4	1	0.0000	0.3900	0.3900	0.3960	09/16/2024	16:24
4	CCB1				1	0.0000	0.0000	0.0000	0.0060	09/16/2024	16:27
5	LB132300BL	50	50		1	0.0000	0.0000	0.0000	0.0060	09/16/2024	16:30
6	LB132300BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.4060	09/16/2024	16:33
7	P3848-19	50	50		1	0.0000	1.3000	1.3000	1.3080	09/16/2024	16:36
8	CCV2			0.4	1	0.0000	0.3800	0.3800	0.3860	09/16/2024	16:39
9	CCB2				1	0.0000	0.0000	0.0000	0.0060	09/16/2024	16:42

LB132300

WORKLIST(Hardcopy Internal Chain)

WorkList Name : RESIDUAL CHLORINE-0

WorkList ID : 183536

Department : Wet-Chemistry

Date : 09-16-2024 14:52:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3848-19	PT-CL-WP	Water	Residual Chlorine	Cool 4 deg C	CHEM02	QA Of	09/03/2024	SM4500 Cl G

Date/Time 09-16-2024 15:00
Raw Sample Received by: NF(WC)
Raw Sample Relinquished by: SJ(QAO)

Date/Time NA
Raw Sample Received by: "
Raw Sample Relinquished by: "