

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB132308

ANALYST: rubina
 SUPERVISOR REVIEW BY: Iwona
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W3105
IODINE SOLUTION .025N 1L	W3114
Starch Solution, 4L	W2977

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB163152BL		1	50	50	2.00	0.00	1.92	1.92	0.08	0.00	0.00	09/06/2024	11:10
2	P3821-01		1	50	50	2.00	0.00	1.90	1.90	0.10	0.02	0.16	09/06/2024	11:13
3	P3821-01DUP		1	50	50	2.00	0.00	1.90	1.90	0.10	0.02	0.16	09/06/2024	11:16
4	P3852-06		1	50	50	2.00	0.00	1.84	1.84	0.16	0.08	0.64	09/06/2024	11:19
5	P3853-01		1	50	50	2.00	0.00	1.90	1.90	0.10	0.02	0.16	09/06/2024	11:22

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume