

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132312

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	09/06/2024	13:00
2	P3850-01		1	14	100.00	>212.0	>212.0	09/06/2024	13:30
3	P3852-06		1	11	100.00	>212.0	>212.0	09/06/2024	14:00
4	P3853-01		1	12	100.00	>212.0	>212.0	09/06/2024	14:30
5	P3853-01DUP		1	12	100.00	>212.0	>212.0	09/06/2024	15:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB132312

WorkList Name : fp-9-6-

WorkList ID : 183215

Department : Wet-Chemistry

Date : 09-06-2024 08:34:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3850-01	50428	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/05/2024	1010B
P3852-06	D3659	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/05/2024	1010B
P3853-01	252805	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/05/2024	1010B

Date/Time 09/06/2024 12:50
Raw Sample Received by: RM
Raw Sample Relinquished by: RM

Date/Time 09/06/2024 15:10
Raw Sample Received by: RM
Raw Sample Relinquished by: RM