

LB132325


SFAM01.1

INSTRUMENT ID: CYL

Linear

 μ Abs.:

37782



A= 0.0000e+000

B= 2.6334e-004 slope

C= 3.7882e-002 y-intercept

Rho= 0.9997615

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.069	0.069	118	0.000	118					1
0.05	0.050					442	0				22
0.20	0.200	0.244	0.044	781	0.0 %	781					5
2.50	2.500	2.370	-0.130	8857	0.0 %	8857					1
5.00	5.000	4.928	-0.072	18571	0.0 %	18571					1
7.50	7.500	7.601	0.101	28720	0.0 %	28720					1
10.0	10.000	9.988	-0.012	37782	0.0 %	37782					0

LB132325

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	118	-		0 SFAM01.1	PPB	9/9/2024 11:28	Std
	0.2 S01	781	-		0.2 SFAM01.1	PPB	9/9/2024 11:31	Std
	2.5 S02	8857	-		2.5 SFAM01.1	PPB	9/9/2024 11:33	Std
	5 S03	18571	-		5 SFAM01.1	PPB	9/9/2024 11:35	Std
	7.5 S04	28720	-		7.5 SFAM01.1	PPB	9/9/2024 11:37	Std
	10 S05	37782	-		10 SFAM01.1	PPB	9/9/2024 11:42	Std
ICV013	ICV013	14901	3.962 -		SFAM01.1	PPB	9/9/2024 11:45	SMPL
ICB013	ICB013	-128	0.0042 -		SFAM01.1	PPB	9/9/2024 11:47	SMPL
CCV045	CCV045	19335	5.1297 -		SFAM01.1	PPB	9/9/2024 11:50	SMPL
CCB045	CCB045	-121	0.006 -		SFAM01.1	PPB	9/9/2024 11:52	SMPL
P3677-05	MX125	16302	4.3309 -		SFAM01.1	PPB	9/9/2024 11:54	SMPL
P3723-29	MH1Z70	-158	-0.0037 -		SFAM01.1	PPB	9/9/2024 11:57	SMPL
P3723-30	MH1Z72	45	0.0497 -		SFAM01.1	PPB	9/9/2024 11:59	SMPL
P3723-31	MH1Z74	74	0.0574 -		SFAM01.1	PPB	9/9/2024 12:01	SMPL
P3723-32	MH1Z76	1123	0.3336 -		SFAM01.1	PPB	9/9/2024 12:03	SMPL
P3723-33	MH1Z78	-21	0.0324 -		SFAM01.1	PPB	9/9/2024 12:06	SMPL
P3723-34	MH1Z80	-33	0.0292 -		SFAM01.1	PPB	9/9/2024 12:08	SMPL
P3723-35	MH1Z82	41	0.0487 -		SFAM01.1	PPB	9/9/2024 12:10	SMPL
P3723-36	MH1Z84	-139	0.0013 -		SFAM01.1	PPB	9/9/2024 12:12	SMPL
P3723-37	MH1Z86	355	0.1314 -		SFAM01.1	PPB	9/9/2024 12:15	SMPL
P3723-38	MH1Z88	-150	-0.0016 -		SFAM01.1	PPB	9/9/2024 12:17	SMPL
P3723-39	MH1Z90	-12	0.0347 -		SFAM01.1	PPB	9/9/2024 12:19	SMPL
P3723-40	MH1Z92	0	0.0379 -		SFAM01.1	PPB	9/9/2024 12:21	SMPL
P3723-41	MH1Z94	-2	0.0374 -		SFAM01.1	PPB	9/9/2024 12:24	SMPL
P3723-42	MH1Z96	10	0.0405 -		SFAM01.1	PPB	9/9/2024 12:26	SMPL
P3723-43	MH1Z97	10	0.0405 -		SFAM01.1	PPB	9/9/2024 12:28	SMPL
P3723-44	MH1Z97D	3	0.0387 -		SFAM01.1	PPB	9/9/2024 12:31	SMPL
P3723-45	MH1Z97S	4166	1.135 -		SFAM01.1	PPB	9/9/2024 12:33	SMPL
P3831-01	MC0AA5	-24	0.0316 -		SFAM01.1	PPB	9/9/2024 12:35	SMPL
P3831-02	MC0AA5D	3	0.0387 -		SFAM01.1	PPB	9/9/2024 12:37	SMPL
CCV046	CCV046	18516	4.914 -		SFAM01.1	PPB	9/9/2024 12:40	SMPL
CCB046	CCB046	-22	0.0321 -		SFAM01.1	PPB	9/9/2024 12:42	SMPL
P3831-03	MC0AA5S	4021	1.0968 -		SFAM01.1	PPB	9/9/2024 12:44	SMPL
P3831-04	MC0AA6	-110	0.0089 -		SFAM01.1	PPB	9/9/2024 12:46	SMPL
PB163221BL	PBW221	-380	-0.0622 -		SFAM01.1	PPB	9/9/2024 12:53	SMPL
CCV047	CCV047	18679	4.9569 -		SFAM01.1	PPB	9/9/2024 12:56	SMPL
CCB047	CCB047	-231	-0.023 -		SFAM01.1	PPB	9/9/2024 12:58	SMPL