

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132333

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	09/09/2024	14:00
2	P3872-02		1	11	100.00	>212.0	>212.0	09/09/2024	15:10
3	P3872-02DUP		1	11	100.00	>212.0	>212.0	09/09/2024	15:40
4	P3889-01		1	14	100.00	>212.0	>212.0	09/09/2024	16:10
5	P3896-01		1	5	10.00	50.0	49.3	09/09/2024	16:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132333

WorkList Name : fp-9-9

WorkList ID : 183276

Department : Wet-Chemistry

Date : 09-09-2024 10:36:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3872-02	WC-W-07-202409	Water	Flash Point	Cool 4 deg C	AECO02	H53	09/05/2024	1010B
P3889-01	FMI109	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/06/2024	1010B
P3896-01	MOO-23-00292	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/06/2024	1010B

Date/Time 09/09/2024 13:50
 Raw Sample Received by: RM cwc
 Raw Sample Relinquished by: RM cwc

Date/Time 09/09/2024 16:50
 Raw Sample Received by: RM cwc
 Raw Sample Relinquished by: RM cwc