

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB132336

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/09/2024	16:47
2	CAL2	1	Water	NA	NA	20.3	7.00	09/09/2024	16:48
3	CAL3	1	Water	NA	NA	20.2	10.02	09/09/2024	16:49
4	ICV	1	Water	NA	NA	20.4	7.01	09/09/2024	16:50
5	CCV1	1	Water	NA	NA	20.4	2.02	09/09/2024	18:10
6	P3872-02	1	Water	NA	NA	20.2	7.13	09/09/2024	18:11
7	P3889-01	1	Water	NA	NA	24.3	7.83	09/09/2024	18:15
8	P3896-01	1	Water	NA	NA	24.2	11.40	09/09/2024	18:20
9	P3912-01	1	Water	NA	NA	20.7	7.43	09/09/2024	18:24
10	P3912-02	1	Water	NA	NA	20.7	7.87	09/09/2024	18:25
11	P3913-04	1	Water	NA	NA	22.1	5.58	09/09/2024	18:29
12	P3913-06	1	Water	NA	NA	22.6	5.51	09/09/2024	18:33
13	P3913-08	1	Water	NA	NA	22.4	5.50	09/09/2024	18:37
14	P3913-10	1	Water	NA	NA	23.2	5.80	09/09/2024	18:40
15	P3913-10DUP	1	Water	NA	NA	23.3	5.81	09/09/2024	18:41
16	CCV2	1	Water	NA	NA	20.3	12.02	09/09/2024	18:44

WORKLIST(Hardcopy Internal Chain)

VB 132336

WorkList Name : ph p3872

WorkList ID : 183282

Department : Wet-Chemistry

Date : 09-09-2024 10:59:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3872-02	WC-W-07-202409	Water	pH	Cool 4 deg C	AECO02	H53	09/05/2024	9040C
P3889-01	FMI109	Water	pH	Cool 4 deg C	PSEG03	H53	09/06/2024	9040C
P3896-01	MOO-23-00292	Water	pH	Cool 4 deg C	PSEG03	H53	09/06/2024	9040C
P3912-01	282-COMPOSITE-1	Water	pH	Cool 4 deg C	PSEG03	H53	09/09/2024	9040C
P3912-02	270-COMPOSITE-1	Water	pH	Cool 4 deg C	PSEG03	H53	09/09/2024	9040C
P3913-04	9424-A-C	Water	pH	Cool 4 deg C	PSEG03	H53	09/09/2024	9040C
P3913-06	9424-D-F	Water	pH	Cool 4 deg C	PSEG03	H53	09/09/2024	9040C
P3913-08	9424-D-F-82924-D	Water	pH	Cool 4 deg C	PSEG03	H53	09/09/2024	9040C
P3913-10	82924-A-C	Water	pH	Cool 4 deg C	PSEG03	H53	09/09/2024	9040C

Date/Time 09.09.24 16:45

Raw Sample Received by: Jd wcc

Raw Sample Relinquished by: Jd wcc

Date/Time 09.09.24

Raw Sample Received by: Jd wcc

Raw Sample Relinquished by: Jd wcc

19:00

Jd wcc

Jd wcc