

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132341

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	09/10/2024	13:45
2	P3839-05		1	11	100.00	>212.0	>212.0	09/10/2024	14:15
3	P3839-05DUP		1	11	100.00	>212.0	>212.0	09/10/2024	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132341

WorkList Name : fp s-9-6

WorkList ID : 183227

Department : Wet-Chemistry

Date : 09-06-2024 11:05:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3839-05	822	Solid	Flash Point	Cool 4 deg C	PSEG03	H53	09/04/2024	1010B

Date/Time 09/10/2024 13:35
 Raw Sample Received by: RY CWL
 Raw Sample Relinquished by: Acco

Date/Time 09/10/2024 15:20
 Raw Sample Received by: RY CWL
 Raw Sample Relinquished by: RY CWL