


LB132312

SFAM01.1

INSTRUMENT ID: CV1



A= 0.0000e+000
 B= 2.6528e-004 slope
 C= 2.1652e-002 y-intercept
 Rho= 0.9990502
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OP
0.00	0.000	0.021	0.021	-4	0.000	-4					1
0.05	0.050					442	0				2
0.20	0.200	0.204	0.004	688	0.0 %	688					4
2.50	2.500	2.410	-0.090	9003	0.0 %	9003					1
5.00	5.000	4.958	-0.042	18609	0.0 %	18609					4
7.50	7.500	7.817	0.317	29387	0.0 %	29387					2
10.0	10.000	9.790	-0.210	36823	0.0 %	36823					

LB132352

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-4	-		0 SFAM01.1	PPB	9/10/2024 14:33	Std
	0.2 S01	688	-		0.2 SFAM01.1	PPB	9/10/2024 14:35	Std
	2.5 S02	9003	-		2.5 SFAM01.1	PPB	9/10/2024 14:37	Std
	5 S03	18609	-		5 SFAM01.1	PPB	9/10/2024 14:40	Std
	7.5 S04	29387	-		7.5 SFAM01.1	PPB	9/10/2024 14:42	Std
	10 S05	36823	-		10 SFAM01.1	PPB	9/10/2024 14:47	Std
ICV015	ICV015	14709	3.9236 -		SFAM01.1	PPB	9/10/2024 14:50	SMPL
ICB015	ICB015	-251	-0.0449 -		SFAM01.1	PPB	9/10/2024 14:52	SMPL
CCV052	CCV052	19144	5.1001 -		SFAM01.1	PPB	9/10/2024 14:54	SMPL
CCB052	CCB052	-266	-0.0489 -		SFAM01.1	PPB	9/10/2024 14:57	SMPL
PB163269BL	PBW269	-17	0.0171 -		SFAM01.1	PPB	9/10/2024 14:59	SMPL
P3730-01	MH1Z71	11	0.0246 -		SFAM01.1	PPB	9/10/2024 15:01	SMPL
P3730-02	MH1Z73	38	0.0317 -		SFAM01.1	PPB	9/10/2024 15:03	SMPL
P3730-03	MH1Z75	60	0.0376 -		SFAM01.1	PPB	9/10/2024 15:06	SMPL
P3730-04	MH1Z77	1474	0.4127 -		SFAM01.1	PPB	9/10/2024 15:08	SMPL
P3730-05	MH1Z79	4	0.0227 -		SFAM01.1	PPB	9/10/2024 15:10	SMPL
P3730-06	MH1Z81	42	0.0328 -		SFAM01.1	PPB	9/10/2024 15:13	SMPL
P3730-07	MH1Z83	84	0.0439 -		SFAM01.1	PPB	9/10/2024 15:15	SMPL
P3730-08	MH1Z85	-64	0.0047 -		SFAM01.1	PPB	9/10/2024 15:17	SMPL
P3730-09	MH1Z87	197	0.0739 -		SFAM01.1	PPB	9/10/2024 15:19	SMPL
P3730-10	MH1Z89	-18	0.0169 -		SFAM01.1	PPB	9/10/2024 15:22	SMPL
P3730-11	MH1Z91	48	0.0344 -		SFAM01.1	PPB	9/10/2024 15:24	SMPL
P3730-12	MH1Z93	54	0.036 -		SFAM01.1	PPB	9/10/2024 15:26	SMPL
P3730-13	MH1Z95	69	0.04 -		SFAM01.1	PPB	9/10/2024 15:28	SMPL
P3730-14	MH1Z95D	86	0.0445 -		SFAM01.1	PPB	9/10/2024 15:31	SMPL
P3730-15	MH1Z95S	4086	1.1056 -		SFAM01.1	PPB	9/10/2024 15:33	SMPL
PB163270BL	PBW270	-17	0.0171 -		SFAM01.1	PPB	9/10/2024 15:35	SMPL
P3756-01	MDCUW6	25	0.0283 -		SFAM01.1	PPB	9/10/2024 15:37	SMPL
P3756-02	MDCUW8	42	0.0328 -		SFAM01.1	PPB	9/10/2024 15:40	SMPL
P3756-03	MDCUX1	1875	0.519 -		SFAM01.1	PPB	9/10/2024 15:42	SMPL
CCV053	CCV053	17669	4.7088 -		SFAM01.1	PPB	9/10/2024 15:44	SMPL
CCB053	CCB053	-190	-0.0287 -		SFAM01.1	PPB	9/10/2024 15:47	SMPL
P3756-04	MDCUX4	1124	0.3198 -		SFAM01.1	PPB	9/10/2024 15:49	SMPL
P3756-05	MDCUX7	-79	0.0007 -		SFAM01.1	PPB	9/10/2024 15:51	SMPL
P3756-06	MDCUX7D	22	0.0275 -		SFAM01.1	PPB	9/10/2024 15:53	SMPL
P3756-07	MDCUX7S	3734	1.0122 -		SFAM01.1	PPB	9/10/2024 15:56	SMPL
P3756-08	MDCVH5	8	0.0238 -		SFAM01.1	PPB	9/10/2024 15:58	SMPL
P3756-09	MDCVN2	166	0.0657 -		SFAM01.1	PPB	9/10/2024 16:00	SMPL
P3756-10	MDCVP9	31	0.0299 -		SFAM01.1	PPB	9/10/2024 16:02	SMPL
P3756-11	MDCVQ3	38	0.0317 -		SFAM01.1	PPB	9/10/2024 16:05	SMPL
P3757-01	MDCUW7	15	0.0256 -		SFAM01.1	PPB	9/10/2024 16:07	SMPL
P3757-02	MDCUX0	50	0.0349 -		SFAM01.1	PPB	9/10/2024 16:09	SMPL
P3757-03	MDCUX3	1113	0.3169 -		SFAM01.1	PPB	9/10/2024 16:11	SMPL
P3757-04	MDCUX6	231	0.0829 -		SFAM01.1	PPB	9/10/2024 16:14	SMPL
P3757-05	MDCUX9	-16	0.0174 -		SFAM01.1	PPB	9/10/2024 16:16	SMPL
P3757-06	MDCUX9D	47	0.0341 -		SFAM01.1	PPB	9/10/2024 16:18	SMPL

LB132352

INSTRUMENT ID : CV1

P3757-07	MDCUX9S	4301	1.1626 -	SFAM01.1	PPB	9/10/2024 16:21 SMPL
P3757-08	MDCVH6	-7	0.0198 -	SFAM01.1	PPB	9/10/2024 16:23 SMPL
P3757-09	MDCVN3	132	0.0567 -	SFAM01.1	PPB	9/10/2024 16:25 SMPL
P3757-10	MDCVP8	38	0.0317 -	SFAM01.1	PPB	9/10/2024 16:27 SMPL
P3757-11	MDCVQ0	32	0.0301 -	SFAM01.1	PPB	9/10/2024 16:30 SMPL
CCV054	CCV054	17117	4.5624 -	SFAM01.1	PPB	9/10/2024 16:32 SMPL
CCB054	CCB054	-145	-0.0168 -	SFAM01.1	PPB	9/10/2024 16:34 SMPL