



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/11/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:45
In Date: 09/10/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:14
Out Date: 09/11/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132357

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P3921-01	YE0P2	1	1.15	8.30	9.45	9.05	95.2	
P3921-02	YE0P3	2	1.15	8.44	9.59	9.11	94.3	
P3921-03	YE0P4	3	1.15	8.36	9.51	9.05	94.5	
P3921-04	YE0P8	4	1.15	8.48	9.63	9.06	93.3	
P3921-05	YE0P9	5	1.14	8.40	9.54	9.27	96.8	
P3921-06	YE0Q1	6	1.18	8.57	9.75	9.44	96.4	
P3921-07	YE0Q1D	7	1.18	8.57	9.75	9.44	96.4	
P3921-08	YE0Q1S	8	1.18	8.57	9.75	9.44	96.4	
P3921-09	YE0Q7	9	1.11	8.61	9.72	8.89	90.4	
P3921-10	YE0Q8	10	1.16	8.68	9.84	9.00	90.3	
P3921-11	YE0R0	11	1.16	8.50	9.66	8.93	91.4	
P3921-12	YE0S1	12	1.16	8.48	9.64	9.12	93.9	
P3921-13	YE0S3	13	1.16	8.78	9.94	9.45	94.4	
P3921-15	YE0P1	14	1.14	8.71	9.85	9.57	96.8	
P3921-16	YE0P5	15	1.18	8.45	9.63	9.55	99.1	
P3921-17	YE0P6	16	1.15	8.80	9.95	9.86	99.0	
P3921-18	YE0Q2	17	1.18	8.26	9.44	9.26	97.8	
P3921-19	YE0Q3	18	1.19	8.56	9.75	9.54	97.5	
P3921-20	YE0Q4	19	1.15	8.68	9.83	9.54	96.7	
P3921-21	YE0Q5	20	1.15	8.55	9.7	9.31	95.4	
P3921-22	YE0Q6	21	1.18	8.51	9.69	9.21	94.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$