

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB132362

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/10/2024	16:25
2	CAL2	1	Water	NA	NA	20.3	7.00	09/10/2024	16:26
3	CAL3	1	Water	NA	NA	20.3	10.02	09/10/2024	16:30
4	ICV	1	Water	NA	NA	20.3	7.00	09/10/2024	16:33
5	CCV1	1	Water	NA	NA	20.3	2.02	09/10/2024	16:50
6	P3926-01	1	Water	NA	NA	20.2	6.94	09/10/2024	17:00
7	P3926-01DUP	1	Water	NA	NA	20.3	6.95	09/10/2024	17:01
8	CCV2	1	Water	NA	NA	20.3	12.02	09/10/2024	17:10

WORKLIST(Hardcopy Internal Chain)

VB 132362

WorkList Name : ph p3926 WorkList ID : 183353 Department : Wet-Chemistry Date : 09-10-2024 16:04:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3926-01	DRAIN WATER TANK-1	Water	pH	Cool 4 deg C	MAJO01	I51	09/10/2024	9040C

Date/Time 09.10.24 16.10
Raw Sample Received by: TA Carr
Raw Sample Relinquished by: [Signature]

Date/Time 09.10.24 17.30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]