

LB132366

7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

3602E

A= 0.0000e+000

B= 2.7805e-004

C= 6.0466e-002

Rho= 0.9997325

Accept=Accepted

Max. Conc.: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	0.061	0.061	3	0.000	3	0				
0.2	0.200	0.228	0.028	601	0.0 %	601					14
2.5	2.500	2.506	0.006	8796	0.0 %	8796					0
5.0	5.000	4.818	-0.182	17111	0.0 %	17111					14
7.5	7.500	7.509	0.009	26790	0.0 %	26790					0
10.0	10.000	10.077	0.077	36026	0.0 %	36026					1

LB132366 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Sep 2024 08:51:31

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	3	-	0.0000	7470A	PPB	11 Sep 2024 08:51:58	S	Std
0.2 - 1	50.2	601	-	0.2000	7470A	PPB	11 Sep 2024 08:54:15	S	Std
2.5 - 1	52.5	8796	-	2.5000	7470A	PPB	11 Sep 2024 08:56:32	S	Std
5.0 - 1	55	17111	-	5.0000	7470A	PPB	11 Sep 2024 08:58:49	S	Std
7.5 - 1	57.5	26790	-	7.5000	7470A	PPB	11 Sep 2024 09:03:44	S	Std
10.0 - 1	59	36026	-	10.0000	7470A	PPB	11 Sep 2024 09:06:01	S	Std
ICV17 - 1	ICV17	14545	4.1047	-	7470A	PPB	11 Sep 2024 09:08:50	U	SMPL
ICB17 - 1	ICB17	-224	-0.0018	-	7470A	PPB	11 Sep 2024 09:11:05	U	SMPL
CCV55 - 1	CCV55	17489	4.9232	-	7470A	PPB	11 Sep 2024 09:13:22	U	SMPL
CCB55 - 1	CCB55	-298	-0.0224	-	7470A	PPB	11 Sep 2024 09:15:37	U	SMPL
CRA - 1	CRA	579	0.2215	-	7470A	PPB	11 Sep 2024 09:17:54	U	SMPL
HighStd - 1	HighStd	35775	10.0076	-	7470A	PPB	11 Sep 2024 09:20:09	U	SMPL
ChkStd - 1	ChkStd	23558	6.6107	-	7470A	PPB	11 Sep 2024 09:22:24	U	SMPL
PB163282BL - 1	PB163282BL	-261	-0.0121	-	7470A	PPB	11 Sep 2024 09:24:42	U	SMPL
PB163282BS - 1	PB163282BS	12576	3.5572	-	7470A	PPB	11 Sep 2024 09:26:59	U	SMPL
P3880-02 - 1	SHEEPSHEAD-SOIL-PILE	-103	0.0318	-	7470A	PPB	11 Sep 2024 09:29:15	U	SMPL
P3881-02 - 1	VNJ-217	-15	0.0563	-	7470A	PPB	11 Sep 2024 09:31:34	U	SMPL
P3891-02 - 1	AUD-SBG	3890	1.1421	-	7470A	PPB	11 Sep 2024 09:33:51	U	SMPL
P3892-02 - 1	ARS20-005	-285	-0.0188	-	7470A	PPB	11 Sep 2024 09:36:07	U	SMPL
P3894-01 - 1	SHOP-DRUM	-25	0.0535	-	7470A	PPB	11 Sep 2024 09:38:25	U	SMPL
CCV56 - 1	CCV56	16695	4.7025	-	7470A	PPB	11 Sep 2024 09:40:40	U	SMPL
CCB56 - 1	CCB56	-87	0.0363	-	7470A	PPB	11 Sep 2024 09:42:55	U	SMPL
P3896-02 - 1	MOO-23-00293	84	0.0838	-	7470A	PPB	11 Sep 2024 09:45:12	U	SMPL
P3905-06 - 1	COMP-1	78	0.0822	-	7470A	PPB	11 Sep 2024 09:47:27	U	SMPL
P3906-02 - 1	SU-701-COMP-01	26	0.0677	-	7470A	PPB	11 Sep 2024 09:49:42	U	SMPL
P3906-08 - 1	SU-701-COMP-02	-24	0.0538	-	7470A	PPB	11 Sep 2024 09:51:57	U	SMPL
P3906-14 - 1	SU-701-COMP-03	2	0.0610	-	7470A	PPB	11 Sep 2024 09:54:13	U	SMPL
P3907-02 - 1	PL-701-COMP-54	0	0.0605	-	7470A	PPB	11 Sep 2024 09:56:29	U	SMPL
P3907-08 - 1	PL-701-COMP-55	11	0.0635	-	7470A	PPB	11 Sep 2024 09:58:46	U	SMPL
P3909-02 - 1	SOM-24-00099	54	0.0755	-	7470A	PPB	11 Sep 2024 10:01:03	U	SMPL
P3909-03 - 1	COMP-1	89	0.0852	-	7470A	PPB	11 Sep 2024 10:03:20	U	SMPL
P3909-04 - 1	COMP-2	377	0.1653	-	7470A	PPB	11 Sep 2024 10:05:36	U	SMPL
CCV57 - 1	CCV57	16688	4.7005	-	7470A	PPB	11 Sep 2024 10:07:51	U	SMPL
CCB57 - 1	CCB57	-235	-0.0049	-	7470A	PPB	11 Sep 2024 10:10:06	U	SMPL
P3911-02 - 1	SP-1	333	0.1531	-	7470A	PPB	11 Sep 2024 10:12:24	U	SMPL
P3911-04 - 1	SP-2	491	0.1970	-	7470A	PPB	11 Sep 2024 10:14:39	U	SMPL
P3911-06 - 1	SP-3	177	0.1097	-	7470A	PPB	11 Sep 2024 10:16:54	U	SMPL
P3911-06DUP - 1	SP-3DUP	150	0.1022	-	7470A	PPB	11 Sep 2024 10:19:10	U	SMPL
P3911-06MS - 1	SP-3MS	17011	4.7903	-	7470A	PPB	11 Sep 2024 10:21:26	U	SMPL
P3911-06MSD - 1	SP-3MSD	15930	4.4898	-	7470A	PPB	11 Sep 2024 10:23:42	U	SMPL
PB163283BL - 1	PBW	-90	0.0354	-	7470A	PPB	11 Sep 2024 10:26:00	U	SMPL
PB163283BS - 1	LCSW	12753	3.6064	-	7470A	PPB	11 Sep 2024 10:28:18	U	SMPL
P3872-02 - 1	WC-W-07-202409	-85	0.0368	-	7470A	PPB	11 Sep 2024 10:30:35	U	SMPL
P3889-01 - 1	FMI109	-13	0.0569	-	7470A	PPB	11 Sep 2024 10:32:52	U	SMPL
CCV58 - 1	CCV58	17157	4.8309	-	7470A	PPB	11 Sep 2024 10:35:08	U	SMPL
CCB58 - 1	CCB58	-174	0.0121	-	7470A	PPB	11 Sep 2024 10:37:23	U	SMPL
P3896-01 - 1	MOO-23-00292	75	0.0813	-	7470A	PPB	11 Sep 2024 10:39:41	U	SMPL
P3913-05 - 1	9424-A-C	2250	0.6861	-	7470A	PPB	11 Sep 2024 10:41:56	U	SMPL
P3913-07 - 1	9424-D-F	827	0.2904	-	7470A	PPB	11 Sep 2024 10:44:12	U	SMPL
P3913-09 - 1	9424-D-F-82924-D	535	0.2092	-	7470A	PPB	11 Sep 2024 10:46:28	U	SMPL
P3913-11 - 1	82924-A-C	355	0.1592	-	7470A	PPB	11 Sep 2024 10:48:44	U	SMPL
P3913-11DUP - 1	82924-A-CDUP	333	0.1531	-	7470A	PPB	11 Sep 2024 10:51:00	U	SMPL
P3913-11MS - 1	82924-A-CMS	17461	4.9155	-	7470A	PPB	11 Sep 2024 10:53:16	U	SMPL
P3913-11MSD - 1	82924-A-CMSD	17737	4.9922	-	7470A	PPB	11 Sep 2024 10:55:33	U	SMPL
CCV59 - 1	CCV59	16885	4.7553	-	7470A	PPB	11 Sep 2024 10:57:51	U	SMPL
CCB59 - 1	CCB59	-112	0.0293	-	7470A	PPB	11 Sep 2024 11:00:09	U	SMPL
PB163203TB - 1		100	0.0883	-	7470A	PPB	11 Sep 2024 11:02:26	U	SMPL
PB163205TB - 1		91	0.0858	-	7470A	PPB	11 Sep 2024 11:04:41	U	SMPL
P3911-06LX5 - 1		40	0.0716	-	7470A	PPB	11 Sep 2024 11:06:57	U	SMPL
P3911-06A - 1		16605	4.6775	-	7470A	PPB	11 Sep 2024 11:09:12	U	SMPL
P3913-11LX5 - 1		-93	0.0346	-	7470A	PPB	11 Sep 2024 11:11:28	U	SMPL
P3913-11A - 1		17910	5.0403	-	7470A	PPB	11 Sep 2024 11:13:45	U	SMPL
CCV60 - 1	CCV60	18742	5.2716	-	7470A	PPB	11 Sep 2024 11:23:03	U	SMPL
CCB60 - 1	CCB60	-62	0.0432	-	7470A	PPB	11 Sep 2024 11:25:19	U	SMPL