

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB132374

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	09/11/2024	11:20
2	P3909-01		1	6	61.00	141.8	141.8	09/11/2024	11:50
3	P3912-01		1	12	100.00	>212.0	>212.0	09/11/2024	12:20
4	P3912-01DUP		1	12	100.00	>212.0	>212.0	09/11/2024	12:50
5	P3912-02		1	11	100.00	>212.0	>212.0	09/11/2024	13:20
6	P3913-04		1	12	100.00	>212.0	>212.0	09/11/2024	13:50
7	P3913-06		1	12	100.00	>212.0	>212.0	09/11/2024	14:20
8	P3913-08		1	11	100.00	>212.0	>212.0	09/11/2024	14:50
9	P3913-10		1	12	100.00	>212.0	>212.0	09/11/2024	15:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132374

WorkList Name : FP9-10 WorkList ID : 183333 Department : Wet-Chemistry Date : 09-10-2024 10:25:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3909-01	SOM-24-00101	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/06/2024	1010B
P3912-01	282-COMPOSITE-1	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/09/2024	1010B
P3912-02	270-COMPOSITE-1	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/09/2024	1010B
P3913-04	9424-A-C	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/09/2024	1010B
P3913-06	9424-D-F	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/09/2024	1010B
P3913-08	9424-D-F-82924-D	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/09/2024	1010B
P3913-10	82924-A-C	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/09/2024	1010B

Date/Time 09/11/24 10:30
 Raw Sample Received by: 12600
 Raw Sample Relinquished by: 48600

Date/Time 09/11/24 16:00
 Raw Sample Received by: 12600
 Raw Sample Relinquished by: 12600