

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132376

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/11/2024	14:45
2	CAL2	1	Water	NA	NA	20.3	7.00	09/11/2024	14:46
3	CAL3	1	Water	NA	NA	20.3	10.02	09/11/2024	14:47
4	ICV	1	Water	NA	NA	20.3	7.02	09/11/2024	14:50
5	CCV1	1	Water	NA	NA	20.3	2.01	09/11/2024	14:55
6	P3948-02	1	Water	NA	NA	22.3	5.55	09/11/2024	15:00
7	P3953-01	1	Water	NA	NA	20.1	6.54	09/11/2024	15:10
8	P3955-04	1	Water	NA	NA	24.7	11.75	09/11/2024	15:25
9	P3955-04DUP	1	Water	NA	NA	24.8	11.76	09/11/2024	15:26
10	CCV2	1	Water	NA	NA	20.3	12.03	09/11/2024	15:30

WORKLIST(Hardcopy Internal Chain)

UB 132376

WorkList Name : ph p3953

WorkList ID : 183390

Department : Wet-Chemistry

Date : 09-11-2024 14:32:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3948-02	C 50413	Water	pH	Cool 4 deg C	PSEG03	H53	09/11/2024	9040C
P3953-01	B DRAIN WATER TANK-1	Water	pH	Cool 4 deg C	MAJO01	I51	09/11/2024	9040C
P3955-04	A 291551	Water	pH	Cool 4 deg C	PSEG03	H53	09/11/2024	9040C

Date/Time 09-11-24 16:40

Raw Sample Received by: JH Celso

Raw Sample Relinquished by: RM SM

Date/Time 09-11-24 16:40

Raw Sample Received by: RM SM

Raw Sample Relinquished by: JP aWC