

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132377

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/11/2024	14:45
2	CAL2	1	Water	NA	NA	20.3	7.00	09/11/2024	14:46
3	CAL3	1	Water	NA	NA	20.3	10.02	09/11/2024	14:47
4	ICV	1	Water	NA	NA	20.3	7.02	09/11/2024	14:50
5	CCV1	1	Water	NA	NA	20.4	2.02	09/11/2024	15:35
6	P3948-01	1	Solid	20.02	20	22.4	5.90	09/11/2024	15:40
7	P3948-01DUP	1	Solid	20.03	20	22.5	5.91	09/11/2024	15:41
8	CCV2	1	Water	NA	NA	20.4	12.02	09/11/2024	15:45

WORKLIST(Hardcopy Internal Chain)

LB 132377

WorkList Name : ph s p3948

WorkList ID : 183391

Department : Wet-Chemistry

Date : 09-11-2024 14:34:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3948-01 C	50228	Solid	pH	Cool 4 deg C	PSEG03	H53	09/11/2024	9045D

Date/Time 09-11-24 14:40
Raw Sample Received by: Jd wdc, RM SM
Raw Sample Relinquished by: RM SM

Date/Time 09-11-25 16:40
Raw Sample Received by: RM SM
Raw Sample Relinquished by: Jd wdc