

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB132378

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/11/2024	13:15
2	CAL2	1	Water	NA	NA	20.3	7.00	09/11/2024	13:16
3	CAL3	1	Water	NA	NA	20.3	10.02	09/11/2024	13:19
4	ICV	1	Water	NA	NA	20.3	7.02	09/11/2024	13:20
5	CCV1	1	Water	NA	NA	20.2	2.02	09/11/2024	13:25
6	P3923-01	1	Solid	20.02	20	23.6	6.40	09/11/2024	13:30
7	P3923-01DUP	1	Solid	20.03	20	23.7	6.41	09/11/2024	13:31
8	P3936-01	1	Solid	20.02	20	22.0	6.09	09/11/2024	13:35
9	P3947-04	1	Solid	20.03	20	21.9	6.62	09/11/2024	13:40
10	P3954-01	1	Solid	20.02	20	24.3	5.67	09/11/2024	13:50
11	CCV2	1	Water	NA	NA	20.2	12.02	09/11/2024	13:55