

LB132387


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39502



A= 0.0000e+000

B= 2.5524e-004 *slope*C= 2.6212e-002 *=intercept*

Rho= 0.9998118

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.049	0.049	90	0.000	90					5
0.05	0.050					442	0				8
0.20	0.200	0.215	0.015	739	0.0 %	739					-1
2.50	2.500	2.481	-0.019	9618	0.0 %	9618					-1
5.00	5.000	4.968	-0.032	19361	0.0 %	19361					-2
7.50	7.500	7.378	-0.122	28803	0.0 %	28803					1
10.0	10.000	10.109	0.109	39502	0.0 %	39502					

LB132387

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	90	-		0 SFAM01.1	PPB	9/11/2024 18:12	Std
	0.2 S01	739	-		0.2 SFAM01.1	PPB	9/11/2024 18:14	Std
	2.5 S02	9618	-		2.5 SFAM01.1	PPB	9/11/2024 18:17	Std
	5 S03	19361	-		5 SFAM01.1	PPB	9/11/2024 18:19	Std
	7.5 S04	28803	-		7.5 SFAM01.1	PPB	9/11/2024 18:21	Std
	10 S05	39502	-		10 SFAM01.1	PPB	9/11/2024 18:24	Std
ICV019	ICV019	15469	3.9746 -		SFAM01.1	PPB	9/11/2024 18:27	SMPL
ICB019	ICB019	-53	0.0127 -		SFAM01.1	PPB	9/11/2024 18:29	SMPL
CCV067	CCV067	19221	4.9323 -		SFAM01.1	PPB	9/11/2024 18:31	SMPL
CCB067	CCB067	-78	0.0063 -		SFAM01.1	PPB	9/11/2024 18:33	SMPL
PB163315BL	PBS315	-12	0.0231 -		SFAM01.1	PPB	9/11/2024 18:36	SMPL
P3920-01	E1KL1	5734	1.4898 -		SFAM01.1	PPB	9/11/2024 18:38	SMPL
P3920-02	E1KL3	1353	0.3716 -		SFAM01.1	PPB	9/11/2024 18:40	SMPL
P3920-03	E1KL6	1454	0.3973 -		SFAM01.1	PPB	9/11/2024 18:43	SMPL
P3920-04	E1KL7	258	0.0921 -		SFAM01.1	PPB	9/11/2024 18:45	SMPL
P3920-05	E1KL9	154	0.0655 -		SFAM01.1	PPB	9/11/2024 18:47	SMPL
P3920-06	E1KM1	207	0.079 -		SFAM01.1	PPB	9/11/2024 18:49	SMPL
P3920-07	E1KM5	1075	0.3006 -		SFAM01.1	PPB	9/11/2024 18:52	SMPL
P3920-08	E1KM7	68	0.0436 -		SFAM01.1	PPB	9/11/2024 18:54	SMPL
P3920-09	E1KM8	127	0.0586 -		SFAM01.1	PPB	9/11/2024 18:56	SMPL
P3920-10	E1KN1	118	0.0563 -		SFAM01.1	PPB	9/11/2024 18:58	SMPL
P3920-11	E1KN3	134	0.0604 -		SFAM01.1	PPB	9/11/2024 19:01	SMPL
P3920-12	E1KN4	484	0.1497 -		SFAM01.1	PPB	9/11/2024 19:03	SMPL
P3920-13	E1KN6	202	0.0778 -		SFAM01.1	PPB	9/11/2024 19:05	SMPL
P3920-14	E1KN8	100	0.0517 -		SFAM01.1	PPB	9/11/2024 19:07	SMPL
P3920-15	E1KP0	109	0.054 -		SFAM01.1	PPB	9/11/2024 19:10	SMPL
P3920-16	E1KP2	104	0.0528 -		SFAM01.1	PPB	9/11/2024 19:12	SMPL
P3920-17	E1KQ0	1009	0.2838 -		SFAM01.1	PPB	9/11/2024 19:14	SMPL
P3920-18	EIKQ2	1060	0.2968 -		SFAM01.1	PPB	9/11/2024 19:17	SMPL
P3920-19	EIKQ2D	992	0.2794 -		SFAM01.1	PPB	9/11/2024 19:19	SMPL
CCV068	CCV068	19792	5.078 -		SFAM01.1	PPB	9/11/2024 19:21	SMPL
CCB068	CCB068	-88	0.0038 -		SFAM01.1	PPB	9/11/2024 19:23	SMPL
P3920-20	EIKQ2S	11044	2.8451 -		SFAM01.1	PPB	9/11/2024 19:26	SMPL
P3920-21	EIKQ4	10	0.0288 -		SFAM01.1	PPB	9/11/2024 19:28	SMPL
P3920-22	EIKQ6	127	0.0586 -		SFAM01.1	PPB	9/11/2024 19:30	SMPL
PB163316BL	PBS316	-2	0.0257 -		SFAM01.1	PPB	9/11/2024 19:32	SMPL
P3924-01	EIKQ8	192	0.0752 -		SFAM01.1	PPB	9/11/2024 19:35	SMPL
P3924-02	EIKQ8D	177	0.0714 -		SFAM01.1	PPB	9/11/2024 19:37	SMPL
P3924-03	EIKQ8S	10985	2.8301 -		SFAM01.1	PPB	9/11/2024 19:39	SMPL
CCV069	CCV069	19568	5.0208 -		SFAM01.1	PPB	9/11/2024 19:41	SMPL
CCB069	CCB069	-94	0.0022 -		SFAM01.1	PPB	9/11/2024 19:44	SMPL