

LB132395


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44249



A= 0.0000e+000

B= 2.2488e-004 *slope*C= 4.7351e-002 *intercept*

Rho= 0.9999170

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%R)
0.00	0.000	0.040	0.040	-32	0.000	-32					
0.05	0.050					442	0				
0.20	0.200	0.221	0.021	770	0.0 %	770					
2.50	2.500	2.398	-0.102	10451	0.0 %	10451					
5.00	5.000	5.028	0.028	22148	0.0 %	22148					
7.50	7.500	7.516	0.016	33210	0.0 %	33210					
10.0	10.000	9.998	-0.002	44249	0.0 %	44249					

0/01
=
11
-4
1
0
0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-32	-		0 SFAM01.1	PPB	9/12/2024 10:12	Std
	0.2 S01	770	-		0.2 SFAM01.1	PPB	9/12/2024 10:14	Std
	2.5 S02	10451	-		2.5 SFAM01.1	PPB	9/12/2024 10:17	Std
	5 S03	22148	-		5 SFAM01.1	PPB	9/12/2024 10:19	Std
	7.5 S04	33210	-		7.5 SFAM01.1	PPB	9/12/2024 10:21	Std
	10 S05	44249	-		10 SFAM01.1	PPB	9/12/2024 10:23	Std
ICV020	ICV020	16919	3.8521 -		SFAM01.1	PPB	9/12/2024 10:30	SMPL
ICB020	ICB020	-227	-0.0037 -		SFAM01.1	PPB	9/12/2024 10:32	SMPL
CCV070	CCV070	21719	4.9315 -		SFAM01.1	PPB	9/12/2024 10:34	SMPL
CCB070	CCB070	-271	-0.0136 -		SFAM01.1	PPB	9/12/2024 10:37	SMPL
PB163332BL	PBW332	-36	0.0393 -		SFAM01.1	PPB	9/12/2024 10:39	SMPL
P3724-01	MH1Z51	95	0.0687 -		SFAM01.1	PPB	9/12/2024 10:41	SMPL
P3724-02	MH1Z52	102	0.0703 -		SFAM01.1	PPB	9/12/2024 10:43	SMPL
P3724-03	MH1Z53	122	0.0748 -		SFAM01.1	PPB	9/12/2024 10:46	SMPL
P3724-05	MH1Z55	-36	0.0393 -		SFAM01.1	PPB	9/12/2024 10:48	SMPL
P3724-06	MH1Z56	-57	0.0345 -		SFAM01.1	PPB	9/12/2024 10:50	SMPL
P3724-07	MH1Z57	-56	0.0348 -		SFAM01.1	PPB	9/12/2024 10:53	SMPL
P3724-08	MH1Z58	-38	0.0388 -		SFAM01.1	PPB	9/12/2024 10:55	SMPL
P3724-09	MH1Z59	-27	0.0413 -		SFAM01.1	PPB	9/12/2024 10:57	SMPL
P3724-10	MH1Z60	23	0.0525 -		SFAM01.1	PPB	9/12/2024 10:59	SMPL
P3724-11	MH1Z61	32	0.0545 -		SFAM01.1	PPB	9/12/2024 11:02	SMPL
P3724-12	MH1Z62	48	0.0581 -		SFAM01.1	PPB	9/12/2024 11:04	SMPL
P3724-13	MH1Z63	126	0.0757 -		SFAM01.1	PPB	9/12/2024 11:06	SMPL
P3724-14	MH1Z64	100	0.0698 -		SFAM01.1	PPB	9/12/2024 11:08	SMPL
P3724-15	MH1Z65	20	0.0518 -		SFAM01.1	PPB	9/12/2024 11:11	SMPL
P3724-16	MH1Z66	48	0.0581 -		SFAM01.1	PPB	9/12/2024 11:13	SMPL
P3724-17	MH1Z67	79	0.0651 -		SFAM01.1	PPB	9/12/2024 11:15	SMPL
P3724-18	MH1Z68	61	0.0611 -		SFAM01.1	PPB	9/12/2024 11:17	SMPL
P3724-19	MH1Z69	37	0.0557 -		SFAM01.1	PPB	9/12/2024 11:20	SMPL
P3724-20	MH1Z54	34	0.055 -		SFAM01.1	PPB	9/12/2024 11:22	SMPL
CCV071	CCV071	20666	4.6947 -		SFAM01.1	PPB	9/12/2024 11:24	SMPL
CCB071	CCB071	-212	-0.0003 -		SFAM01.1	PPB	9/12/2024 11:26	SMPL
P3724-21	MH1Z54D	77	0.0647 -		SFAM01.1	PPB	9/12/2024 11:29	SMPL
P3724-22	MH1Z54S	4897	1.1486 -		SFAM01.1	PPB	9/12/2024 11:31	SMPL
PB163333BL	PBW333	-38	0.0388 -		SFAM01.1	PPB	9/12/2024 11:33	SMPL
P3806-01	MDCVM6	595	0.1812 -		SFAM01.1	PPB	9/12/2024 11:36	SMPL
P3806-02	MDCVN0	1330	0.3464 -		SFAM01.1	PPB	9/12/2024 11:38	SMPL
P3806-03	MDCVN0D	1343	0.3494 -		SFAM01.1	PPB	9/12/2024 11:40	SMPL
P3806-04	MDCVN0S	5429	1.2682 -		SFAM01.1	PPB	9/12/2024 11:42	SMPL
P3806-05	MDCVM4	-84	0.0285 -		SFAM01.1	PPB	9/12/2024 11:45	SMPL
P3807-01	MDCVM7	168	0.0851 -		SFAM01.1	PPB	9/12/2024 11:47	SMPL
P3807-02	MDCVN1	59	0.0606 -		SFAM01.1	PPB	9/12/2024 11:49	SMPL
P3807-03	MDCVN1D	111	0.0723 -		SFAM01.1	PPB	9/12/2024 11:51	SMPL
P3807-04	MDCVN1S	4476	1.0539 -		SFAM01.1	PPB	9/12/2024 11:54	SMPL
P3807-05	MDCVM5	-4	0.0465 -		SFAM01.1	PPB	9/12/2024 11:56	SMPL
CCV072	CCV072	19956	4.5351 -		SFAM01.1	PPB	9/12/2024 11:58	SMPL

LB132395

INSTRUMENT ID : CV1

CCB072

CCB072

-250 -0.0089 -

SFAM01.1 PPB

9/12/2024 12:00 SMPL