

LB132401


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

39510



A= 0.0000e+000

B= 2.5315e-004

C= 2.4951e-004

Rho= 0.9997858

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/v)
0.0	0.000	0.062	0.062	243	0.000	243	0				12
0.2	0.200	0.224	0.024	882	0.0 %	882					2
2.5	2.500	2.448	-0.052	9671	0.0 %	9671					3
5.0	5.000	4.865	-0.135	19218	0.0 %	19218					1
7.5	7.500	7.599	0.099	30017	0.0 %	30017					0
10.0	10.000	10.002	0.002	39510	0.0 %	39510					

LB32401 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Sep 2024 12:29:30

Sample ID	Extended ID	Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	243	-	0.0000	7470A	PPB	12 Sep 2024 12:33:18	S	Std
0.2 - 1	50.2	882	-	0.2000	7470A	PPB	12 Sep 2024 12:35:35	S	Std
2.5 - 1	12.5	9671	-	2.5000	7470A	PPB	12 Sep 2024 12:37:52	S	Std
5.0 - 1	55	19218	-	5.0000	7470A	PPB	12 Sep 2024 12:40:09	S	Std
7.5 - 1	55.5	30017	-	7.5000	7470A	PPB	12 Sep 2024 12:47:52	S	Std
10.0 - 1	59.5	39510	-	10.0000	7470A	PPB	12 Sep 2024 12:52:43	S	Std
ICV21 - 1	ICV21	15118	3.8273	-	7470A	PPB	12 Sep 2024 12:55:27	U	SMPL
ICB21 - 1	ICB21	-12	-0.0028	-	7470A	PPB	12 Sep 2024 12:57:43	U	SMPL
CCV73 - 1	CCV73	18985	4.8062	-	7470A	PPB	12 Sep 2024 13:00:00	U	SMPL
CCB73 - 1	CCB73	-1	-0.0000	-	7470A	PPB	12 Sep 2024 13:02:15	U	SMPL
CRA - 1	CRA	835	0.2116	-	7470A	PPB	12 Sep 2024 13:04:31	U	SMPL
HighStd - 1	HighStd	39297	9.9482	-	7470A	PPB	12 Sep 2024 13:06:47	U	SMPL
ChkStd - 1	ChkStd	27144	6.8717	-	7470A	PPB	12 Sep 2024 13:09:02	U	SMPL
PB163327BL - 1	PBW	-196	-0.0494	-	7470A	PPB	12 Sep 2024 13:11:20	U	SMPL
PB163327BS - 1	LCSW	18120	4.5873	-	7470A	PPB	12 Sep 2024 13:13:39	U	SMPL
P3912-01 - 1	282-COMPOSITE-1	15678	3.9691	-	7470A	PPB	12 Sep 2024 13:15:55	U	SMPL
P3912-02 - 1	270-COMPOSITE-1	4963	1.2566	-	7470A	PPB	12 Sep 2024 13:18:13	U	SMPL
P3913-04 - 1	9424-A-C	1239	0.3139	-	7470A	PPB	12 Sep 2024 13:20:32	U	SMPL
P3913-06 - 1	9424-D-F	827	0.2096	-	7470A	PPB	12 Sep 2024 13:22:50	U	SMPL
P3913-08 - 1	9424-D-F-82924-D	1056	0.2676	-	7470A	PPB	12 Sep 2024 13:25:06	U	SMPL
CCV74 - 1	CCV74	20921	5.2963	-	7470A	PPB	12 Sep 2024 13:27:21	U	SMPL
CCB74 - 1	CCB74	187	0.0476	-	7470A	PPB	12 Sep 2024 13:29:36	U	SMPL
P3913-10 - 1	82924-A-C	1357	0.3438	-	7470A	PPB	12 Sep 2024 13:31:52	U	SMPL
P3955-04 - 1	291551	107	0.0273	-	7470A	PPB	12 Sep 2024 13:34:08	U	SMPL
P3961-01 - 1	1807	746	0.1891	-	7470A	PPB	12 Sep 2024 13:36:24	U	SMPL
P3961-01DUP - 1	1807DUP	699	0.1772	-	7470A	PPB	12 Sep 2024 13:38:40	U	SMPL
P3961-01MS - 1	1807MS	11344	2.8719	-	7470A	PPB	12 Sep 2024 13:40:56	U	SMPL
P3961-01MSD - 1	1807MSD	10240	2.5925	-	7470A	PPB	12 Sep 2024 13:43:12	U	SMPL
PB163224BL - 1	PBW	23	0.0061	-	7470A	PPB	12 Sep 2024 13:48:09	U	SMPL
PB163224BS - 1	LCSW	16039	4.0605	-	7470A	PPB	12 Sep 2024 13:50:25	U	SMPL
P3852-06 - 1	D3659	946	0.2397	-	7470A	PPB	12 Sep 2024 13:52:42	U	SMPL
P3852-06DUP - 1	D3659DUP	925	0.2344	-	7470A	PPB	12 Sep 2024 13:57:45	U	SMPL
CCV75 - 1	CCV75	20885	5.2872	-	7470A	PPB	12 Sep 2024 14:00:01	U	SMPL
CCB75 - 1	CCB75	132	0.0337	-	7470A	PPB	12 Sep 2024 14:02:16	U	SMPL
P3852-06MS - 1	D3659MS	11978	3.0324	-	7470A	PPB	12 Sep 2024 14:04:34	U	SMPL
P3852-06MSD - 1	D3659MSD	12277	3.1081	-	7470A	PPB	12 Sep 2024 14:06:49	U	SMPL
P3857-11 - 1		95	0.0243	-	7470A	PPB	12 Sep 2024 14:09:07	U	SMPL
P3857-11DUP - 1		213	0.0542	-	7470A	PPB	12 Sep 2024 14:11:24	U	SMPL
P3857-11MS - 1		15474	3.9174	-	7470A	PPB	12 Sep 2024 14:13:40	U	SMPL
P3857-11MSD - 1		15849	4.0124	-	7470A	PPB	12 Sep 2024 14:15:56	U	SMPL
P3857-11LX5 - 1		-53	-0.0132	-	7470A	PPB	12 Sep 2024 14:18:14	U	SMPL
P3857-11A - 1		15033	3.8058	-	7470A	PPB	12 Sep 2024 14:20:32	U	SMPL
PB163170TB - 1	PB163170TB	-17	-0.0041	-	7470A	PPB	12 Sep 2024 14:22:49	U	SMPL
PB163171TB - 1	PB163171TB	80	0.0205	-	7470A	PPB	12 Sep 2024 14:25:06	U	SMPL
CCV76 - 1	CCV76	19678	4.9817	-	7470A	PPB	12 Sep 2024 14:27:21	U	SMPL
CCB76 - 1	CCB76	25	0.0066	-	7470A	PPB	12 Sep 2024 14:29:36	U	SMPL
P3961-01LX5 - 1		29	0.0076	-	7470A	PPB	12 Sep 2024 14:31:53	U	SMPL
P3961-01A - 1		16123	4.0817	-	7470A	PPB	12 Sep 2024 14:34:09	U	SMPL
P3852-06LX5 - 1		112	0.0286	-	7470A	PPB	12 Sep 2024 14:36:24	U	SMPL
P3852-06A - 1		16092	4.0739	-	7470A	PPB	12 Sep 2024 14:38:41	U	SMPL
CCV77 - 1	CCV77	19781	5.0077	-	7470A	PPB	12 Sep 2024 14:40:58	U	SMPL
CCB77 - 1	CCB77	-2	-0.0003	-	7470A	PPB	12 Sep 2024 14:43:15	U	SMPL