

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132402

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	09/12/2024	12:00
2	P3948-02		1	12	100.00	>212.0	>212.0	09/12/2024	13:15
3	P3955-04		1	14	100.00	>212.0	>212.0	09/12/2024	13:45
4	P3961-01		1	11	100.00	>212.0	>212.0	09/12/2024	14:15
5	P3961-01DUP		1	11	100.00	>212.0	>212.0	09/12/2024	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66132402

WorkList Name : FP9-11 WorkList ID : 183389 Department : Wet-Chemistry Date : 09-11-2024 14:05:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3948-02	50413	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/11/2024	1010B
P3955-04	291551	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/11/2024	1010B
P3961-01	1807	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/11/2024	1010B

Date/Time 09/12/2024 11:50
Raw Sample Received by: RM CWG
Raw Sample Relinquished by: RM CWG

Date/Time 09/12/2024 15:20
Raw Sample Received by: RM CWG
Raw Sample Relinquished by: RM CWG