


LB132408

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.020	-0.020	302	0.000	302					
0.2	0.200	0.205	0.005	1399	0.0 %	1399					
2.5	2.500	2.505	0.005	12620	0.0 %	12620					
5.0	5.000	5.008	0.008	24829	0.0 %	24829					
7.5	7.500	7.530	0.030	37130	0.0 %	37130					
10.0	10.000	9.972	-0.028	49044	0.0 %	49044					

0.00021

LB132408 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Sep 2024 15:03:31

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	302	-	0.0000	7471B	PPB	12 Sep 2024 15:35:32	S	Std
0.2 - 1	502	1399	-	0.2000	7471B	PPB	12 Sep 2024 15:37:49	S	Std
2.5 - 1	52.5	12620	-	2.5000	7471B	PPB	12 Sep 2024 15:40:06	S	Std
5.0 - 1	55	24829	-	5.0000	7471B	PPB	12 Sep 2024 15:42:23	S	Std
7.5 - 1	57.5	37130	-	7.5000	7471B	PPB	12 Sep 2024 15:47:29	S	Std
10.0 - 1	510	49044	-	10.0000	7471B	PPB	12 Sep 2024 15:52:28	S	Std
ICV22 - 1	ICV22	21587	4.3434	-	7471B	PPB	12 Sep 2024 15:57:53	U	SMPL
ICB22 - 1	ICB22	151	-0.0509	-	7471B	PPB	12 Sep 2024 16:00:09	U	SMPL
CCV78 - 1	CCV78	25010	5.0451	-	7471B	PPB	12 Sep 2024 16:02:28	U	SMPL
CCB78 - 1	CCB78	164	-0.0483	-	7471B	PPB	12 Sep 2024 16:04:42	U	SMPL
CRA - 1	CRA	1433	0.2119	-	7471B	PPB	12 Sep 2024 16:07:00	U	SMPL
HighStd - 1	HighStd	49137	9.9911	-	7471B	PPB	12 Sep 2024 16:09:15	U	SMPL
ChkStd - 1	ChkStd	34882	7.0689	-	7471B	PPB	12 Sep 2024 16:11:30	U	SMPL
PB163329BL - 1	PBS	255	-0.0296	-	7471B	PPB	12 Sep 2024 16:16:38	U	SMPL
PB163329BS - 1	LCSS	19556	3.9271	-	7471B	PPB	12 Sep 2024 16:18:53	U	SMPL
P3925-01 - 1	NB-07091024	3480	0.6315	-	7471B	PPB	12 Sep 2024 16:21:09	U	SMPL
P3925-01DUP - 1	NB-07091024DUP	3011	0.5354	-	7471B	PPB	12 Sep 2024 16:23:29	U	SMPL
P3925-01MS - 1	NB-07091024MS	24599	4.9609	-	7471B	PPB	12 Sep 2024 16:25:48	U	SMPL
P3925-01MSD - 1	NB-07091024MSD	22460	4.5224	-	7471B	PPB	12 Sep 2024 16:28:07	U	SMPL
P3929-01 - 1	BIN-2	2759	0.4837	-	7471B	PPB	12 Sep 2024 16:30:25	U	SMPL
CCV79 - 1	CCV79	24247	4.8887	-	7471B	PPB	12 Sep 2024 16:32:44	U	SMPL
CCB79 - 1	CCB79	216	-0.0376	-	7471B	PPB	12 Sep 2024 16:35:01	U	SMPL
P3929-03 - 1	72-11930	3592	0.6545	-	7471B	PPB	12 Sep 2024 16:37:19	U	SMPL
P3930-01 - 1	DN-B-33	3745	0.6858	-	7471B	PPB	12 Sep 2024 16:39:33	U	SMPL
P3930-02 - 1	DN-B-34	11047	2.1827	-	7471B	PPB	12 Sep 2024 16:41:50	U	SMPL
P3936-01 - 1	DRUM-COMP	2302	0.3900	-	7471B	PPB	12 Sep 2024 16:44:07	U	SMPL
P3947-01 - 1	RW-1	3206	0.5754	-	7471B	PPB	12 Sep 2024 16:46:25	U	SMPL
P3949-01 - 1	BREAKER-2546-BASE	2022	0.3326	-	7471B	PPB	12 Sep 2024 16:48:43	U	SMPL
P3962-01 - 1	TR-04-091124	2673	0.4661	-	7471B	PPB	12 Sep 2024 16:51:00	U	SMPL
P3963-01 - 1	DN-B-39	4482	0.8369	-	7471B	PPB	12 Sep 2024 16:53:17	U	SMPL
P3925-01LX5 - 1		222	-0.0364	-	7471B	PPB	12 Sep 2024 16:55:35	U	SMPL
P3925-01A - 1		23332	4.7012	-	7471B	PPB	12 Sep 2024 16:57:52	U	SMPL
CCV80 - 1	CCV80	25008	5.0447	-	7471B	PPB	12 Sep 2024 17:00:07	U	SMPL
CCB80 - 1	CCB80	284	-0.0237	-	7471B	PPB	12 Sep 2024 17:02:26	U	SMPL