

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB132409

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/12/2024	15:25
2	CAL2	1	Water	NA	NA	20.3	7.00	09/12/2024	15:26
3	CAL3	1	Water	NA	NA	20.3	10.02	09/12/2024	15:30
4	ICV	1	Water	NA	NA	20.2	7.01	09/12/2024	15:31
5	CCV1	1	Water	NA	NA	20.3	2.01	09/12/2024	15:35
6	P3961-01	1	Water	NA	NA	22.3	5.85	09/12/2024	15:40
7	P3977-01	1	Water	NA	NA	20.3	7.33	09/12/2024	15:45
8	P3977-01DUP	1	Water	NA	NA	20.4	7.34	09/12/2024	15:46
9	CCV2	1	Water	NA	NA	20.3	12.02	09/12/2024	15:50

WORKLIST(Hardcopy Internal Chain)

132409

WorkList Name : ph p3962 WorkList ID : 183441 Department : Wet-Chemistry Date : 09-12-2024 15:17:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3961-01	1807	Water	pH	Cool 4 deg C	PSEG03	H53	09/11/2024	9040C
P3977-01	DRAIN-WATER-TANK-1	Water	pH	Cool 4 deg C	MAJO01	H53	09/12/2024	9040C

Date/Time 09/12/24 15:20
Raw Sample Received by: JS (EXA-1066)
Raw Sample Relinquished by: JS (EXA-1066)

Date/Time 09-12-24 18:00
Raw Sample Received by: JS (EXA-1066)
Raw Sample Relinquished by: JS (EXA-1066)