

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132410

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/12/2024	15:25
2	CAL2	1	Water	NA	NA	20.3	7.00	09/12/2024	15:26
3	CAL3	1	Water	NA	NA	20.3	10.02	09/12/2024	15:30
4	ICV	1	Water	NA	NA	20.2	7.01	09/12/2024	15:35
5	CCV1	1	Water	NA	NA	20.2	2.02	09/12/2024	16:00
6	P3984-01	1	Solid	20.03	20	24.7	8.33	09/12/2024	16:14
7	P3984-04	1	Solid	20.04	20	24.9	8.44	09/12/2024	16:22
8	P3984-07	1	Solid	20.01	20	24.7	8.33	09/12/2024	16:27
9	P3984-07DUP	1	Solid	20.02	20	24.8	8.34	09/12/2024	16:28
10	CCV2	1	Water	NA	NA	20.2	12.02	09/12/2024	16:33

132410

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p3984 WorkList ID : 183442 Department : Wet-Chemistry Date : 09-12-2024 15:33:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3984-01	WC-A-COMP	Solid	pH	Cool 4 deg C	EART12	H53	09/12/2024	9045D
P3984-04	WC-B-COMP	Solid	pH	Cool 4 deg C	EART12	H53	09/12/2024	9045D
P3984-07	WC-C-COMP	Solid	pH	Cool 4 deg C	EART12	H53	09/12/2024	9045D

Date/Time 09.12.24 15:45

Raw Sample Received by: JO WOC

Raw Sample Relinquished by: RECEIVED-KAW

Date/Time 09.12.24 18:30

Raw Sample Received by: RECEIVED-KAW

Raw Sample Relinquished by: JO WOC