

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB132411

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/12/2024	15:25
2	CAL2	1	Water	NA	NA	20.3	7.00	09/12/2024	15:26
3	CAL3	1	Water	NA	NA	20.3	10.02	09/12/2024	15:30
4	ICV	1	Water	NA	NA	20.2	7.01	09/12/2024	15:35
5	CCV1	1	Water	NA	NA	20.2	2.02	09/12/2024	16:00
6	P3976-04	1	Solid	20.02	20	24.4	7.42	09/12/2024	16:05
7	P3984-02	1	Solid	20.03	20	24.7	8.33	09/12/2024	16:14
8	P3984-05	1	Solid	20.04	20	24.9	8.44	09/12/2024	16:22
9	P3984-08	1	Solid	20.01	20	24.7	8.33	09/12/2024	16:27
10	P3984-08DUP	1	Solid	20.02	20	24.8	8.34	09/12/2024	16:28
11	CCV2	1	Water	NA	NA	20.2	12.02	09/12/2024	16:33

VB 132411

Date: 09-12-2024 15:33:57

Date/Time 09.12.24
Raw Sample Received by: Jol WLC
Raw Sample Relinquished by: RJ C. Hall-Jab

Page 1 of 1