

LB132415


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

37977



A= 0.0000e+000

B= 2.6309e-004 slope

C= 1.2854e-002 y-intercept

Rho= 0.9999898

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	46D
0.00	0.000	-0.000	-0.000	-50	0.000	-50					
0.05	0.050					442	0				
0.20	0.200	0.189	-0.011	671	0.0 %	671					
2.50	2.500	2.504	0.004	9469	0.0 %	9469					
5.00	5.000	5.029	0.029	19065	0.0 %	19065					
7.50	7.500	7.474	-0.026	28358	0.0 %	28358					
10.0	10.000	10.004	0.004	37977	0.0 %	37977					

46D
1160100

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-50	-		0 SFAM01.1	PPB	9/13/2024 11:02	Std
	0.2 S01	671	-		0.2 SFAM01.1	PPB	9/13/2024 11:04	Std
	2.5 S02	9469	-		2.5 SFAM01.1	PPB	9/13/2024 11:06	Std
	5 S03	19065	-		5 SFAM01.1	PPB	9/13/2024 11:08	Std
	7.5 S04	28358	-		7.5 SFAM01.1	PPB	9/13/2024 11:11	Std
	10 S05	37977	-		10 SFAM01.1	PPB	9/13/2024 11:17	Std
ICV023	ICV023	16031	4.2305 -		SFAM01.1	PPB	9/13/2024 11:23	SMPL
ICB023	ICB023	-167	-0.0311 -		SFAM01.1	PPB	9/13/2024 11:25	SMPL
CCV081	CCV081	19812	5.2253 -		SFAM01.1	PPB	9/13/2024 11:27	SMPL
CCB081	CCB081	-207	-0.0416 -		SFAM01.1	PPB	9/13/2024 11:29	SMPL
PB163379BL	PBS379	-9	0.0105 -		SFAM01.1	PPB	9/13/2024 11:32	SMPL
P3938-01	EIKR2	338	0.1018 -		SFAM01.1	PPB	9/13/2024 11:34	SMPL
P3938-02	EIKR2D	552	0.1581 -		SFAM01.1	PPB	9/13/2024 11:36	SMPL
P3938-03	EIKR2S	12390	3.2726 -		SFAM01.1	PPB	9/13/2024 11:41	SMPL
P3938-04	EIKR3	2687	0.7198 -		SFAM01.1	PPB	9/13/2024 11:43	SMPL
P3938-05	EIKR5	1515	0.4114 -		SFAM01.1	PPB	9/13/2024 11:46	SMPL
P3938-06	EIKR8	195	0.0642 -		SFAM01.1	PPB	9/13/2024 11:48	SMPL
P3938-07	EIKS0	147	0.0515 -		SFAM01.1	PPB	9/13/2024 11:50	SMPL
P3938-08	EIKS2	274	0.0849 -		SFAM01.1	PPB	9/13/2024 11:52	SMPL
P3938-09	EIKS6	624	0.177 -		SFAM01.1	PPB	9/13/2024 11:55	SMPL
PB163380BL	PBS380	-45	0.001 -		SFAM01.1	PPB	9/13/2024 11:57	SMPL
P3974-01	EIKS8	1186	0.3249 -		SFAM01.1	PPB	9/13/2024 11:59	SMPL
P3974-02	EIKT0	787	0.2199 -		SFAM01.1	PPB	9/13/2024 12:01	SMPL
P3974-03	EIKT2	1120	0.3075 -		SFAM01.1	PPB	9/13/2024 12:04	SMPL
P3974-04	EIKT4	683	0.1925 -		SFAM01.1	PPB	9/13/2024 12:06	SMPL
P3974-05	EIKT6	739	0.2073 -		SFAM01.1	PPB	9/13/2024 12:08	SMPL
P3974-06	EIKU3	804	0.2244 -		SFAM01.1	PPB	9/13/2024 12:10	SMPL
P3974-07	EIKU5	691	0.1947 -		SFAM01.1	PPB	9/13/2024 12:13	SMPL
P3974-08	EIKU7	196	0.0644 -		SFAM01.1	PPB	9/13/2024 12:15	SMPL
P3974-09	EIKU9	144	0.0507 -		SFAM01.1	PPB	9/13/2024 12:17	SMPL
CCV082	CCV082	19951	5.2618 -		SFAM01.1	PPB	9/13/2024 12:20	SMPL
CCB082	CCB082	-132	-0.0219 -		SFAM01.1	PPB	9/13/2024 12:22	SMPL
P3974-10	EIKV1	310	0.0944 -		SFAM01.1	PPB	9/13/2024 12:24	SMPL
P3974-11	EIKV4	991	0.2736 -		SFAM01.1	PPB	9/13/2024 12:26	SMPL
P3974-12	EIKV6	378	0.1123 -		SFAM01.1	PPB	9/13/2024 12:29	SMPL
P3974-13	EIKV8	356	0.1065 -		SFAM01.1	PPB	9/13/2024 12:31	SMPL
P3974-14	EIKW0	291	0.0894 -		SFAM01.1	PPB	9/13/2024 12:33	SMPL
P3974-15	EIKW2	308	0.0939 -		SFAM01.1	PPB	9/13/2024 12:35	SMPL
P3974-16	EIKW4	159	0.0547 -		SFAM01.1	PPB	9/13/2024 12:38	SMPL
P3974-17	EIKW4D	137	0.0489 -		SFAM01.1	PPB	9/13/2024 12:40	SMPL
P3974-18	EIKW4S	12554	3.3157 -		SFAM01.1	PPB	9/13/2024 12:42	SMPL
CCV083	CCV083	20127	5.3082 -		SFAM01.1	PPB	9/13/2024 12:44	SMPL
CCB083	CCB083	-130	-0.0213 -		SFAM01.1	PPB	9/13/2024 12:47	SMPL