



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/16/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 09/13/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 09/14/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132425

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P3989-01	TP-2	1	1.15	8.78	9.93	9.31	92.9	
P3989-02	TP-2-VOC	2	1.12	8.73	9.85	9.25	93.1	
P3989-03	TP-2-EPH	3	1.15	7.15	8.3	7.24	85.2	
P3989-05	TP-3	4	1.16	8.50	9.66	8.23	83.2	
P3989-06	TP-3-VOC	5	1.17	8.58	9.75	7.99	79.5	
P3989-07	TP-3-EPH	6	1.18	7.31	8.49	7.94	92.5	
P3996-05	SVOC-GPC-BLANK	7	1.00	1.00	2.00	2.00	100.0	
P3996-06	PEST-GPC-BLANK	8	1.00	1.00	2.00	2.00	100.0	
P3996-07	PEST-GPC-BLANK-SPIKE	9	1.00	1.00	2.00	2.00	100.0	
P3996-08	PCB-GPC-BLANK	10	1.00	1.00	2.00	2.00	100.0	
P3996-09	PCB-GPC-BLANK-SPIKE	11	1.00	1.00	2.00	2.00	100.0	
P3996-10	SVOC-GPC2-BLANK	12	1.00	1.00	2.00	2.00	100.0	
P3996-11	PEST-GPC2-BLANK	13	1.00	1.00	2.00	2.00	100.0	
P3996-12	PEST-GPC2-BLANK-SPKIE	14	1.00	1.00	2.00	2.00	100.0	
P3996-13	PCB-GPC2-BLANK	15	1.00	1.00	2.00	2.00	100.0	
P3996-14	PCB-GPC2-BLANK-SPKIE	16	1.00	1.00	2.00	2.00	100.0	
P4000-01	OK-02-09132024	17	1.13	8.43	9.56	9.21	95.8	
P4000-02	OK-02-09132024-E2	18	1.18	8.56	9.74	9.22	93.9	
P4002-01	TP-1	19	1.18	8.51	9.69	8.62	87.4	
P4002-02	TP-1-E2	20	1.12	8.70	9.82	8.78	88.0	
P4002-03	TP-2	21	1.13	8.61	9.74	8.48	85.4	
P4002-04	TP-2-E2	22	1.17	8.40	9.57	8.34	85.4	
P4002-05	TP-4	23	1.19	8.52	9.71	8.79	89.2	
P4002-06	TP-4-E2	24	1.13	8.70	9.83	8.97	90.1	
P4002-07	TP-5	25	1.15	8.80	9.95	7.91	76.8	
P4002-08	TP-5-E2	26	1.15	8.67	9.82	7.83	77.0	
P4009-01	DN-B-34B	27	1.13	8.76	9.89	8.86	88.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$