


LB132429

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.6900e-004

C= 6.1204e-003

Rho= 0.9998594

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	34	0.000	34	0			
0.2	0.200	0.178	-0.022	638	0.0 %	638				
2.5	2.500	2.474	-0.026	9174	0.0 %	9174				
5.0	5.000	4.997	-0.003	18552	0.0 %	18552				
7.5	7.500	7.620	0.120	28306	0.0 %	28306				
10.0	10.000	9.916	-0.084	36840	0.0 %	36840				

LB132429 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Sep 2024 14:04:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	34	-	0.0000	7470A	PPB	13 Sep 2024 14:09:18	S	Std
0.2 - 1	50.2	638	-	0.2000	7470A	PPB	13 Sep 2024 14:11:35	S	Std
2.5 - 1	52.5	9174	-	2.5000	7470A	PPB	13 Sep 2024 14:13:52	S	Std
5.0 - 1	55	18552	-	5.0000	7470A	PPB	13 Sep 2024 14:16:08	S	Std
7.5 - 1	57.5	28306	-	7.5000	7470A	PPB	13 Sep 2024 14:18:25	S	Std
10.0 - 1	57.0	36840	-	10.0000	7470A	PPB	13 Sep 2024 14:20:42	S	Std
ICV24 - 1	ICV24	14686	3.9566	-	7470A	PPB	13 Sep 2024 14:23:55	U	SMPL
ICB24 - 1	ICB24	-243	-0.0592	-	7470A	PPB	13 Sep 2024 14:26:10	U	SMPL
CCV84 - 1	CCV84	18370	4.9476	-	7470A	PPB	13 Sep 2024 14:28:27	U	SMPL
CCB84 - 1	CCB84	-264	-0.0649	-	7470A	PPB	13 Sep 2024 14:30:42	U	SMPL
CRA - 1	CRA	592	0.1654	-	7470A	PPB	13 Sep 2024 14:32:58	U	SMPL
HighStd - 1	HighStd	36818	9.9101	-	7470A	PPB	13 Sep 2024 14:35:14	U	SMPL
ChkStd - 1	ChkStd	26181	7.0488	-	7470A	PPB	13 Sep 2024 14:37:29	U	SMPL
PB163361BL - 1	PBW	-399	-0.1012	-	7470A	PPB	13 Sep 2024 14:39:47	U	SMPL
PB163361BS - 1	LCSW	14235	3.8353	-	7470A	PPB	13 Sep 2024 14:42:04	U	SMPL
P3932-03 - 1	FM1	-253	-0.0619	-	7470A	PPB	13 Sep 2024 14:44:20	U	SMPL
P3947-04 - 1	RW-1	-54	-0.0084	-	7470A	PPB	13 Sep 2024 14:46:38	U	SMPL
P3949-02 - 1	BREAKER-2546-BASE	-38	-0.0041	-	7470A	PPB	13 Sep 2024 14:48:54	U	SMPL
P3949-02DUP - 1	BREAKER-2546-BASEDUP	11	0.0091	-	7470A	PPB	13 Sep 2024 14:51:11	U	SMPL
P3949-02MS - 1	BREAKER-2546-BASEMS	15490	4.1729	-	7470A	PPB	13 Sep 2024 14:53:26	U	SMPL
CCV85 - 1	CCV85	19491	5.2492	-	7470A	PPB	13 Sep 2024 14:55:42	U	SMPL
CCB85 - 1	CCB85	-306	-0.0762	-	7470A	PPB	13 Sep 2024 14:57:59	U	SMPL
P3949-02MSD - 1	BREAKER-2546-BASEMSD	16032	4.3187	-	7470A	PPB	13 Sep 2024 15:00:16	U	SMPL
PB163307TB - 1	PB163307TB	-245	-0.0598	-	7470A	PPB	13 Sep 2024 15:02:31	U	SMPL
P3949-02LX5 - 1		58	0.0217	-	7470A	PPB	13 Sep 2024 15:04:48	U	SMPL
P3949-02A - 1		15455	4.1635	-	7470A	PPB	13 Sep 2024 15:07:04	U	SMPL
CCV86 - 1	CCV86	19920	5.3646	-	7470A	PPB	13 Sep 2024 15:09:19	U	SMPL
CCB86 - 1	CCB86	-297	-0.0738	-	7470A	PPB	13 Sep 2024 15:11:37	U	SMPL