

Analytical Summary Report

Analysis Method: SM2120 B

Parameter: Color

Run Number: LB132434

Reviewed By: Niha

Supervisor Review By: Iwona

pH Meter ID: PH METER-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP109700
Calibration std, 5 color unit	WP109701
Calibration std, 10 color unit	WP109702
Calibration std, 15 color unit	WP109704
Calibration std, 20 color unit	WP109715
Calibration std, 25 color unit	WP109705
Calibration std, 30 color unit	WP109706
Calibration std, 40 color unit	WP109707
Calibration std, 50 color unit	WP109708
Calibration std, 60 color unit	WP109709
Calibration std, 70 color unit	WP109710
Calibration std, 100 color unit	WP109712

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB132434BL	LB132434BL	1	5.04	N/A	0.00	09/13/2024	15:00
2	P3848-27	PT-COL-WP	1	1.54	N/A	40.00	09/13/2024	15:05

WORKLIST(Hardcopy Internal Chain)

LB132434

WorkList Name : COLOR PT-09132024

WorkList ID : 183490

Department : Wet-Chemistry

Date : 09-13-2024 13:48:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3848-27	PT-COL-WP	Water	Color	Cool 4 deg C	CHEM02	QA Of	09/03/2024	SM2120 B

Date/Time 09.13.2024 , 14:00

Raw Sample Received by: NF(wc)

Raw Sample Relinquished by: SJ(QAO)

Date/Time

NA

Raw Sample Received by: "

Raw Sample Relinquished by: "