

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB132438

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/13/2024	11:00
2	CAL2	1	Water	NA	NA	20.3	7.00	09/13/2024	11:05
3	CAL3	1	Water	NA	NA	20.3	10.02	09/13/2024	11:10
4	ICV	1	Water	NA	NA	20.3	7.00	09/13/2024	11:15
5	CCV1	1	Water	NA	NA	20.3	2.01	09/13/2024	11:20
6	P3989-04	1	Solid	20.02	20	22.7	8.61	09/13/2024	11:25
7	P3989-08	1	Solid	20.03	20	22.5	8.70	09/13/2024	11:30
8	P3989-08DUP	1	Solid	20.04	20	22.6	8.71	09/13/2024	11:31
9	CCV2	1	Water	NA	NA	20.3	12.02	09/13/2024	11:35

WORKLIST(Hardcopy Internal Chain)

✓ 132438

WorkList Name : corrosivity p3989 WorkList ID : 183477 Department : Wet-Chemistry Date : 09-13-2024 10:21:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3989-04	TP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	H53	09/13/2024	9045D
P3989-08	TP-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	H53	09/13/2024	9045D

Date/Time 09-13-24 10:30
Raw Sample Received by: SD WCC
Raw Sample Relinquished by: Jmy (Sm)

Date/Time 09-13-24 13:00
Raw Sample Received by: Jmy (Sm)
Raw Sample Relinquished by: SD WCC