


LB132439

SFAM01.1

INSTRUMENT ID: CV1



Linear



A= 0.0000e+000

B= 2.2360e-004 *slope*C= 2.4578e-002 *y-intercept*

Rho= 0.9999515

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>QAD</i>
0.00	0.000	0.050	0.050	112	0.000	112					
0.05	0.050					442	0				
0.20	0.200	0.205	0.005	807	0.0 %	807					
2.50	2.500	2.448	-0.052	10840	0.0 %	10840					
5.00	5.000	4.957	-0.043	22058	0.0 %	22058					
7.50	7.500	7.526	0.026	33551	0.0 %	33551					
10.0	10.000	10.014	0.014	44675	0.0 %	44675					

QAD
 2
 3
 -2
 -1
 0
 0

LB132439

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	112	-		0 SFAM01.1	PPB	9/13/2024 15:47	Std
	0.2 S01	807	-		0.2 SFAM01.1	PPB	9/13/2024 15:49	Std
	2.5 S02	10840	-		2.5 SFAM01.1	PPB	9/13/2024 15:51	Std
	5 S03	22058	-		5 SFAM01.1	PPB	9/13/2024 15:54	Std
	7.5 S04	33551	-		7.5 SFAM01.1	PPB	9/13/2024 15:56	Std
	10 S05	44675	-		10 SFAM01.1	PPB	9/13/2024 15:58	Std
ICV026	ICV026	17091	3.8461 -		SFAM01.1	PPB	9/13/2024 16:01	SMPL
ICB026	ICB026	-77	0.0074 -		SFAM01.1	PPB	9/13/2024 16:03	SMPL
CCV089	CCV089	21856	4.9115 -		SFAM01.1	PPB	9/13/2024 16:05	SMPL
CCB089	CCB089	-88	0.0049 -		SFAM01.1	PPB	9/13/2024 16:08	SMPL
PB163387BL	PBS163387	65	0.0391 -		SFAM01.1	PPB	9/13/2024 16:10	SMPL
P3992-01	EIKW9	5567	1.2693 -		SFAM01.1	PPB	9/13/2024 16:12	SMPL
P3992-02	EIKX1	921	0.2305 -		SFAM01.1	PPB	9/13/2024 16:14	SMPL
P3992-03	EIKX3	106	0.0483 -		SFAM01.1	PPB	9/13/2024 16:17	SMPL
P3992-04	EIKX5	277	0.0865 -		SFAM01.1	PPB	9/13/2024 16:19	SMPL
P3992-05	EIKX7	75	0.0413 -		SFAM01.1	PPB	9/13/2024 16:21	SMPL
P3992-06	EIKX8	273	0.0856 -		SFAM01.1	PPB	9/13/2024 16:24	SMPL
P3992-07	EIKY0	2035	0.4796 -		SFAM01.1	PPB	9/13/2024 16:26	SMPL
P3992-08	EIKY1	1514	0.3631 -		SFAM01.1	PPB	9/13/2024 16:28	SMPL
P3992-09	EIKY2	809	0.2055 -		SFAM01.1	PPB	9/13/2024 16:30	SMPL
P3992-10	EIKY2D	697	0.1804 -		SFAM01.1	PPB	9/13/2024 16:33	SMPL
P3992-11	EIKY2S	11386	2.5705 -		SFAM01.1	PPB	9/13/2024 16:35	SMPL
CCV090	CCV090	21260	4.7783 -		SFAM01.1	PPB	9/13/2024 16:37	SMPL
CCB090	CCB090	-109	0.0002 -		SFAM01.1	PPB	9/13/2024 16:39	SMPL