


LB132448 INSTRUMENT ID: CV1

7471B

Linear

 μ Abs.:

44711



A= 0.0000e+000

B= 2.2715e-004

C= 3.9397e-002

Rho= 0.9989291

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.048	0.048	40	0.000	40				
0.2	0.200	0.206	0.006	733	0.0 %	733				
2.5	2.500	2.500	0.000	10834	0.0 %	10834				
5.0	5.000	5.102	0.102	22288	0.0 %	22288				
7.5	7.500	7.148	-0.352	31295	0.0 %	31295				
10.0	10.000	10.195	0.195	44711	0.0 %	44711				

1/20
100
+ 2
2

LB132448 INSTRUMENT ID: CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Sep 2024 17:43:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	40	-	0.0000	7471B	PPB	14 Sep 2024 18:57:02	S	Std
0.2 - 1	50.2	733	-	0.2000	7471B	PPB	14 Sep 2024 18:59:19	S	Std
2.5 - 1	52.8	10834	-	2.5000	7471B	PPB	14 Sep 2024 19:01:36	S	Std
5.0 - 1	55.0	22288	-	5.0000	7471B	PPB	14 Sep 2024 19:03:52	S	Std
7.5 - 1	57.8	31295	-	7.5000	7471B	PPB	14 Sep 2024 19:16:03	S	Std
10.0 - 1	510.0	44711	-	10.0000	7471B	PPB	14 Sep 2024 19:37:05	S	Std
ICV28 - 1	ICV28	15879	3.6463	-	7471B	PPB	14 Sep 2024 19:39:57	U	SMPL
ICB28 - 1	ICB28	-269	-0.0217	-	7471B	PPB	14 Sep 2024 19:42:12	U	SMPL
CCV93 - 1	CCV93	21433	4.9078	-	7471B	PPB	14 Sep 2024 19:44:30	U	SMPL
CCB93 - 1	CCB93	-297	-0.0281	-	7471B	PPB	14 Sep 2024 19:46:45	U	SMPL
CRA - 1	CRA	712	0.2011	-	7471B	PPB	14 Sep 2024 19:49:03	U	SMPL
HighStd - 1	HighStd	45126	10.2896	-	7471B	PPB	14 Sep 2024 19:51:18	U	SMPL
ChkStd - 1	ChkStd	31701	7.2402	-	7471B	PPB	14 Sep 2024 19:53:33	U	SMPL
P3976-01 - 1	RW-2	3504	0.8353	-	7471B	PPB	14 Sep 2024 20:05:02	U	SMPL
P3979-01 - 1	EO-03-091224	5195	1.2194	-	7471B	PPB	14 Sep 2024 20:07:18	U	SMPL
P3980-01 - 1	HD-02-091224	1987	0.4907	-	7471B	PPB	14 Sep 2024 20:09:37	U	SMPL
P3984-01 - 1	WC-A-COMP	21149	4.8433	-	7471B	PPB	14 Sep 2024 20:11:55	U	SMPL
P3984-04 - 1	WC-B-COMP	16863	3.8698	-	7471B	PPB	14 Sep 2024 20:14:14	U	SMPL
P3984-07 - 1	WC-C-COMP	39572	9.0280	-	7471B	PPB	14 Sep 2024 20:16:34	U	SMPL
P3985-01 - 1	TP-3	959	0.2572	-	7471B	PPB	14 Sep 2024 20:18:53	U	SMPL
CCV94 - 1	CCV94	21692	4.9667	-	7471B	PPB	14 Sep 2024 20:23:56	U	SMPL
CCB94 - 1	CCB94	164	0.0766	-	7471B	PPB	14 Sep 2024 20:26:10	U	SMPL
P3985-03 - 1	TP-7	1562	0.3942	-	7471B	PPB	14 Sep 2024 20:28:27	U	SMPL
P3989-01 - 1	TP-2	11908	2.7443	-	7471B	PPB	14 Sep 2024 20:30:42	U	SMPL
P3989-05 - 1	TP-3	10922	2.5203	-	7471B	PPB	14 Sep 2024 20:32:58	U	SMPL
P4000-01 - 1	OK-02-09132024	1128	0.2956	-	7471B	PPB	14 Sep 2024 20:35:16	U	SMPL
P4002-01 - 1	TP-1	1186	0.3088	-	7471B	PPB	14 Sep 2024 20:37:34	U	SMPL
P4002-03 - 1	TP-2	1473	0.3740	-	7471B	PPB	14 Sep 2024 20:39:50	U	SMPL
P4002-05 - 1	TP-4	1084	0.2856	-	7471B	PPB	14 Sep 2024 20:42:06	U	SMPL
P4002-07 - 1	TP-5	4688	1.1043	-	7471B	PPB	14 Sep 2024 20:44:23	U	SMPL
P4009-01 - 1	DN-B-34B	44287	10.0990	-	7471B	PPB	14 Sep 2024 20:46:40	U	SMPL
CCV95 - 1	CCV95	22268	5.0975	-	7471B	PPB	14 Sep 2024 20:48:58	U	SMPL
CCB95 - 1	CCB95	149	0.0732	-	7471B	PPB	14 Sep 2024 20:51:20	U	SMPL
PB163402BL - 1		314	0.1107	-	7471B	PPB	14 Sep 2024 20:56:23	U	SMPL
PB163402BS - 1		16944	3.8882	-	7471B	PPB	14 Sep 2024 20:58:39	U	SMPL
P3989-05DUP - 1		8336	1.9329	-	7471B	PPB	14 Sep 2024 21:00:54	U	SMPL
P3989-05LX5 - 1		1777	0.4430	-	7471B	PPB	14 Sep 2024 21:03:12	U	SMPL
P3989-05MS - 1		22122	5.0643	-	7471B	PPB	14 Sep 2024 21:05:30	U	SMPL
P3989-05MSD - 1		22724	5.2011	-	7471B	PPB	14 Sep 2024 21:07:45	U	SMPL
P3989-05A - 1		23917	5.4721	-	7471B	PPB	14 Sep 2024 21:10:05	U	SMPL
P4009-01DLX2 - 1		22628	5.1793	-	7471B	PPB	14 Sep 2024 21:12:24	U	SMPL
CCV96 - 1	CCV96	23020	5.2683	-	7471B	PPB	14 Sep 2024 21:17:29	U	SMPL
CCB96 - 1	CCB96	-191	-0.0040	-	7471B	PPB	14 Sep 2024 21:19:48	U	SMPL