

7470A

LB 132449


CVI

09/15/24

Linear

 μ Abs.:

35935



A= 0.0000e+000

B= 2.8262e-004

C= -1.5983e-002

Rho= 0.9996860

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Adj
0.0	0.000	0.033	0.033	172	0.000	172	0				1.0
0.2	0.200	0.232	0.032	877	0.0 %	877					16
2.5	2.500	2.510	0.010	8938	0.0 %	8938					1
5.0	5.000	4.949	-0.051	17569	0.0 %	17569					-1
7.5	7.500	7.336	-0.164	26013	0.0 %	26013					-2
10.0	10.000	10.140	0.140	35935	0.0 %	35935					2

LB132449 INSTRUMENT ID: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Sep 2024 14:49:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1		172	-	0.0000	7470A	PPB	15 Sep 2024 15:22:29	S	Std
0.2 - 1		877	-	0.2000	7470A	PPB	15 Sep 2024 15:24:46	S	Std
2.5 - 1		8938	-	2.5000	7470A	PPB	15 Sep 2024 15:27:03	S	Std
5.0 - 1		17569	-	5.0000	7470A	PPB	15 Sep 2024 15:29:20	S	Std
7.5 - 1		26013	-	7.5000	7470A	PPB	15 Sep 2024 15:40:22	S	Std
10.0 - 1		35935	-	10.0000	7470A	PPB	15 Sep 2024 15:42:38	S	Std
ICV30 - 1	ICV30	13289	3.7398		7470A	PPB	15 Sep 2024 15:47:28	U	SMPL
ICB30 - 1	ICB30	-280	-0.0951		7470A	PPB	15 Sep 2024 15:49:43	U	SMPL
CCV99 - 1	CCV99	18033	5.0806		7470A	PPB	15 Sep 2024 15:51:59	U	SMPL
CCB99 - 1	CCB99	-321	-0.1067		7470A	PPB	15 Sep 2024 15:54:14	U	SMPL
CRA - 1	CRA	801	0.2104		7470A	PPB	15 Sep 2024 15:56:30	U	SMPL
HighStd - 1	HighStd	34374	9.6989		7470A	PPB	15 Sep 2024 16:01:17	U	SMPL
ChkStd - 1	ChkStd	23410	6.6002		7470A	PPB	15 Sep 2024 16:03:34	U	SMPL
PB163416BL - 1	PBW	-239	-0.0835		7470A	PPB	15 Sep 2024 16:05:52	U	SMPL
PB163416BS - 1	LCSW	14708	4.1408		7470A	PPB	15 Sep 2024 16:08:10	U	SMPL
P3976-04 - 1	RW-2	-71	-0.0360		7470A	PPB	15 Sep 2024 16:10:26	U	SMPL
P3984-02 - 1	WC-A-COMP	-11	-0.0191		7470A	PPB	15 Sep 2024 16:12:43	U	SMPL
P3984-05 - 1	WC-B-COMP	30	-0.0075		7470A	PPB	15 Sep 2024 16:15:00	U	SMPL
P3984-08 - 1	WC-C-COMP	78	0.0061		7470A	PPB	15 Sep 2024 16:17:17	U	SMPL
P3989-04 - 1	TP-2	73	0.0046		7470A	PPB	15 Sep 2024 16:19:32	U	SMPL
CCV100 - 1	CCV100	18084	5.0950		7470A	PPB	15 Sep 2024 16:21:46	U	SMPL
CCB100 - 1	CCB100	-175	-0.0654		7470A	PPB	15 Sep 2024 16:24:01	U	SMPL
P3989-08 - 1	TP-3	120	0.0179		7470A	PPB	15 Sep 2024 16:26:18	U	SMPL
P3989-08DUP - 1	TP-3DUP	133	0.0216		7470A	PPB	15 Sep 2024 16:28:33	U	SMPL
P3989-08MS - 1	TP-3MS	17029	4.7968		7470A	PPB	15 Sep 2024 16:30:48	U	SMPL
P3989-08MSD - 1	TP-3MSD	16322	4.5970		7470A	PPB	15 Sep 2024 16:33:04	U	SMPL
P3989-08LX5 - 1	TP-3L	-103	-0.0451		7470A	PPB	15 Sep 2024 16:35:23	U	SMPL
P3989-08A - 1	TP-3A	16869	4.7516		7470A	PPB	15 Sep 2024 16:37:41	U	SMPL
PB163358TB - 1	LEB358	173	0.0329		7470A	PPB	15 Sep 2024 16:39:57	U	SMPL
CCV01 - 1	CCV01	18000	5.0712		7470A	PPB	15 Sep 2024 16:42:16	U	SMPL
CCB01 - 1	CCB01	-65	-0.0344		7470A	PPB	15 Sep 2024 16:44:33	U	SMPL