



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 9/17/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 09/16/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:13
Out Date: 09/17/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132454

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4014-01	OR-03-09162024	1	1.18	8.44	9.62	9.27	95.9	
P4014-02	OR-03-09162024-E2	2	1.18	8.49	9.67	9.17	94.1	
P4015-01	B-1-TANK17	3	1.18	8.55	9.73	8.79	89.0	
P4015-02	B-2-TANK17	4	1.19	8.37	9.56	8.67	89.4	
P4015-03	SW-1-TANK17	5	1.18	8.48	9.66	8.82	90.1	
P4015-04	SW-2-TANK17	6	1.15	8.80	9.95	8.98	89.0	
P4015-05	SW-3-TANK17	7	1.18	8.72	9.9	8.68	86.0	
P4015-06	P-1-TANK17	8	1.13	8.74	9.87	8.74	87.1	
P4015-07	CHA-1-TANK17	9	1.13	8.74	9.87	8.35	82.6	
P4023-01	TP-9	10	1.14	8.84	9.98	9.16	90.7	
P4023-02	TP-9-E2	11	1.18	8.81	9.99	9.19	90.9	
P4023-03	CONCRETE-TP-9	12	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
P4023-04	CONCRETE-TP-9-E2	13	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
P4025-01	40305	15	1.00	1.00	2.00	2.00	100.0	wipe sample
P4025-02	40306	16	1.00	1.00	2.00	2.00	100.0	wipe sample
P4026-01	MR-CAM-40603	17	1.00	1.00	2.00	2.00	100.0	WIPES / DEBRIS
P4027-01	DN-B-49	14	1.13	8.70	9.83	9.21	92.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$