

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132457

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/16/2024	17:00
2	CAL2	1	Water	NA	NA	20.3	7.00	09/16/2024	17:01
3	CAL3	1	Water	NA	NA	20.2	10.02	09/16/2024	17:05
4	ICV	1	Water	NA	NA	20.3	7.02	09/16/2024	17:10
5	CCV1	1	Water	NA	NA	20.3	2.01	09/16/2024	17:15
6	P4026-01	1	Solid	20.02	20	23.3	7.33	09/16/2024	17:20
7	P4026-01DUP	1	Solid	20.03	20	23.4	7.34	09/16/2024	17:21
8	CCV2	1	Water	NA	NA	20.4	12.02	09/16/2024	17:25

WORKLIST(Hardcopy Internal Chain)

VB 132457

WorkList Name : ph solids p4026

WorkList ID : 183540

Department : Wet-Chemistry

Date : 09-16-2024 16:46:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4026-01	MR-CAM-40603	Solid	pH	Cool 4 deg C	PSEG03	H53	09/16/2024	9045D

Date/Time 09-16-24 16:50

Raw Sample Received by: JG 601,

Raw Sample Relinquished by: JG 601

Date/Time 09-16-24

Raw Sample Received by: JG 601

Raw Sample Relinquished by: JG 601