

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132458

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/16/2024	17:00
2	CAL2	1	Water	NA	NA	20.3	7.00	09/16/2024	17:01
3	CAL3	1	Water	NA	NA	20.2	10.02	09/16/2024	17:05
4	ICV	1	Water	NA	NA	20.3	7.02	09/16/2024	17:10
5	CCV1	1	Water	NA	NA	20.3	2.01	09/16/2024	17:30
6	P3998-01	1	Water	NA	NA	20.9	9.32	09/16/2024	17:37
7	P4030-02	1	Water	NA	NA	20.4	12.10	09/16/2024	17:40
8	P4030-02DUP	1	Water	NA	NA	20.5	12.11	09/16/2024	17:42
9	CCV2	1	Water	NA	NA	20.3	12.02	09/16/2024	17:45

WORKLIST(Hardcopy Internal Chain)

✓ 132458

WorkList Name : ph water p4030 WorkList ID : 183538 Department : Wet-Chemistry Date : 09-16-2024 16:45:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3998-01	DRAIN-WATER-TANK-1	Water	pH	Cool 4 deg C	MAJO01	H53	09/13/2024	9040C
P4030-02	WC-A-08-202409	Water	pH	Cool 4 deg C	AECO02	H53	09/16/2024	9040C

Date/Time 09-16-24 16:50
Raw Sample Received by: JQ WOC
Raw Sample Relinquished by: JQ CSW

Date/Time 09-16-24 18:30
Raw Sample Received by: JQ CSW
Raw Sample Relinquished by: JQ WOC